

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092121\
 Data File : VX024388.D
 Acq On : 21 Sep 2021 16:01
 Operator : JC/MD
 Sample : M3782-19DL 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP04-20210915DL

Quant Time: Sep 21 16:41:11 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 21 10:52:48 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	141655	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	251526	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	241278	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	95579	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	108499	53.714	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	107.420%
35) Dibromofluoromethane	5.397	113	81388	48.249	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.500%
50) Toluene-d8	8.653	98	279161	45.428	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	90.860%#
62) 4-Bromofluorobenzene	11.085	95	110890	46.369	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.740%
Target Compounds						
44) Trichloroethene	7.129	130	66616	35.747	ug/l	Qvalue 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092121\
Data File : VX024388.D
Acq On : 21 Sep 2021 16:01
Operator : JC/MD
Sample : M3782-19DL 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP04-20210915DL

Quant Time: Sep 21 16:41:11 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 21 10:52:48 2021
Response via : Initial Calibration

