

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092121\
 Data File : VX024392.D
 Acq On : 21 Sep 2021 17:36
 Operator : JC/MD
 Sample : M3835-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SW-CB-02-(0.2)

Quant Time: Sep 22 04:51:52 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 21 10:52:48 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	154822	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	228556	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	223875	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	97120	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	114408	51.822	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	103.640%
35) Dibromofluoromethane	5.397	113	91533	59.716	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	119.440%
50) Toluene-d8	8.653	98	281406	50.396	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.800%
62) 4-Bromofluorobenzene	11.085	95	103405	47.585	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.160%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	1809	1.922	ug/l #	79
30) Chloroform	5.093	83	1327	0.417	ug/l #	76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092121\
Data File : VX024392.D
Acq On : 21 Sep 2021 17:36
Operator : JC/MD
Sample : M3835-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SW-CB-02-(0.2)

Quant Time: Sep 22 04:51:52 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 21 10:52:48 2021
Response via : Initial Calibration

