

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092121\
 Data File : VX024393.D
 Acq On : 21 Sep 2021 17:59
 Operator : JC/MD
 Sample : M3835-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SW-RR-01-(1.5)

Quant Time: Sep 22 04:52:04 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 21 10:52:48 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	145989	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	257194	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	259074	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	108061	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	112533	54.057	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 108.120%	
35) Dibromofluoromethane	5.397	113	82980	48.108	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 96.220%	
50) Toluene-d8	8.653	98	322604	51.341	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 102.680%	
62) 4-Bromofluorobenzene	11.085	95	123271	50.410	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 100.820%	
Target Compounds						
16) Acetone	2.392	43	1430	1.611	ug/l	Qvalue 89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092121\
Data File : VX024393.D
Acq On : 21 Sep 2021 17:59
Operator : JC/MD
Sample : M3835-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SW-RR-01-(1.5)

Quant Time: Sep 22 04:52:04 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 21 10:52:48 2021
Response via : Initial Calibration

