

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092121\
 Data File : VX024394.D
 Acq On : 21 Sep 2021 18:23
 Operator : JC/MD
 Sample : M3835-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TRIP-BLANK

Quant Time: Sep 22 04:52:15 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 21 10:52:48 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	145197	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	229357	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	218977	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	89105	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	107979	52.152	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	104.300%
35) Dibromofluoromethane	5.397	113	90534	58.858	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	117.720%
50) Toluene-d8	8.653	98	331882	59.228	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	118.460%#
62) 4-Bromofluorobenzene	11.085	95	111468	51.116	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.240%
Target Compounds						
						Qvalue
16) Acetone	2.392	43	1324	1.500	ug/l	97
20) Methylene Chloride	2.794	84	730	0.431	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092121\
Data File : VX024394.D
Acq On : 21 Sep 2021 18:23
Operator : JC/MD
Sample : M3835-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TRIP-BLANK

Quant Time: Sep 22 04:52:15 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 21 10:52:48 2021
Response via : Initial Calibration

