

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092121\
 Data File : VX024398.D
 Acq On : 21 Sep 2021 19:57
 Operator : JC/MD
 Sample : M3818-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-TT-VPB-RW7-EB-20210915

Quant Time: Sep 22 04:53:00 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 21 10:52:48 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	149577	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.775	114	264252	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	226468	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	103148	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	112763	52.868	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	105.740%
35) Dibromofluoromethane	5.397	113	88849	50.135	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.280%
50) Toluene-d8	8.653	98	290922	45.062	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	90.120%#
62) 4-Bromofluorobenzene	11.085	95	116035	46.184	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.360%
Target Compounds						
						Qvalue
11) Tert butyl alcohol	2.971	59	10559	2.560	ug/l	99
16) Acetone	2.386	43	18060	19.858	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092121\
Data File : VX024398.D
Acq On : 21 Sep 2021 19:57
Operator : JC/MD
Sample : M3818-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-TT-VPB-RW7-EB-20210915

Quant Time: Sep 22 04:53:00 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 21 10:52:48 2021
Response via : Initial Calibration

