

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092121\
 Data File : VX024406.D
 Acq On : 22 Sep 2021 00:16
 Operator : JC/MD
 Sample : M3818-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-TT-VPB-RW7-GW-910-912

Quant Time: Sep 22 04:55:10 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 21 10:52:48 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	131345	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	222962	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	215767	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	90572	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	95909	51.208	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	102.420%
35) Dibromofluoromethane	5.397	113	76087	50.885	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.760%
50) Toluene-d8	8.653	98	273588	50.225	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.460%
62) 4-Bromofluorobenzene	11.085	95	99325	46.854	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.700%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	7629	9.553	ug/l #	84
17) Carbon Disulfide	2.520	76	2224	0.733	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092121\
Data File : VX024406.D
Acq On : 22 Sep 2021 00:16
Operator : JC/MD
Sample : M3818-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-TT-VPB-RW7-GW-910-912

Quant Time: Sep 22 04:55:10 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 21 10:52:48 2021
Response via : Initial Calibration

