

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092121\
 Data File : VX024407.D
 Acq On : 22 Sep 2021 00:40
 Operator : JC/MD
 Sample : M3833-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB03-20210919

Quant Time: Sep 22 04:55:21 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 21 10:52:48 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	127613	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	216922	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	207337	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	86075	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	94696	52.039	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	104.080%
35) Dibromofluoromethane	5.397	113	74902	51.487	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.980%
50) Toluene-d8	8.653	98	266347	50.257	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.520%
62) 4-Bromofluorobenzene	11.085	95	95545	46.326	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.660%
Target Compounds						
16) Acetone	2.392	43	1956	2.521	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092121\
Data File : VX024407.D
Acq On : 22 Sep 2021 00:40
Operator : JC/MD
Sample : M3833-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB03-20210919

Quant Time: Sep 22 04:55:21 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 21 10:52:48 2021
Response via : Initial Calibration

