

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092121\
 Data File : VX024408.D
 Acq On : 22 Sep 2021 01:03
 Operator : JC/MD
 Sample : M3833-19
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EB02-20210920

Quant Time: Sep 22 04:55:33 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 21 10:52:48 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	133026	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	228294	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	218026	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	88098	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	98791	52.080	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	104.160%
35) Dibromofluoromethane	5.397	113	76984	50.282	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.560%
50) Toluene-d8	8.653	98	282039	50.567	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.140%
62) 4-Bromofluorobenzene	11.085	95	99670	45.919	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.840%
Target Compounds						
						Qvalue
11) Tert butyl alcohol	2.977	59	13819	14.642	ug/l	99
16) Acetone	2.386	43	18329	22.661	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092121\
Data File : VX024408.D
Acq On : 22 Sep 2021 01:03
Operator : JC/MD
Sample : M3833-19
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EB02-20210920

Quant Time: Sep 22 04:55:33 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 21 10:52:48 2021
Response via : Initial Calibration

