

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092121\
 Data File : VX024409.D
 Acq On : 22 Sep 2021 01:27
 Operator : JC/MD
 Sample : M3833-20
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB04-20210920

Quant Time: Sep 22 04:55:43 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 21 10:52:48 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	128573	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.775	114	220450	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	215078	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	93715	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	96285	52.517	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	=	105.040%	
35) Dibromofluoromethane	5.403	113	74680	50.513	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	101.020%	
50) Toluene-d8	8.653	98	272257	50.550	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	=	101.100%	
62) 4-Bromofluorobenzene	11.085	95	100312	47.859	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	95.720%	
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	1992	2.548	ug/l	99
20) Methylene Chloride	2.800	84	1210	0.806	ug/l #	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092121\
 Data File : VX024409.D
 Acq On : 22 Sep 2021 01:27
 Operator : JC/MD
 Sample : M3833-20
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_X
 Client Sampled :
 TB04-20210920

Quant Time: Sep 22 04:55:43 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 21 10:52:48 2021
 Response via : Initial Calibration

