

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092121\
 Data File : VX024411.D
 Acq On : 22 Sep 2021 02:14
 Operator : JC/MD
 Sample : M3833-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE123D3-20210918

Quant Time: Sep 22 04:56:06 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 21 10:52:48 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	128782	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	219630	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	210285	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	90321	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	96009	52.281	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 104.560%	
35) Dibromofluoromethane	5.397	113	74026	50.258	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.520%	
50) Toluene-d8	8.653	98	266139	49.599	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 99.200%	
62) 4-Bromofluorobenzene	11.085	95	96734	46.324	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 92.640%	
Target Compounds						
						Qvalue
16) Acetone	2.386	43	1722	2.199	ug/l	# 87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092121\
Data File : VX024411.D
Acq On : 22 Sep 2021 02:14
Operator : JC/MD
Sample : M3833-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE123D3-20210918

Quant Time: Sep 22 04:56:06 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 21 10:52:48 2021
Response via : Initial Calibration

