

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092121\
 Data File : VX024412.D
 Acq On : 22 Sep 2021 02:38
 Operator : JC/MD
 Sample : M3833-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW202D1-20210919

Quant Time: Sep 22 04:56:19 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 21 10:52:48 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	129144	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	223182	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	216855	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	92423	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	99896	54.246	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	108.500%
35) Dibromofluoromethane	5.397	113	76261	50.951	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.900%
50) Toluene-d8	8.652	98	275747	50.572	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.140%
62) 4-Bromofluorobenzene	11.085	95	101537	47.850	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.700%
Target Compounds						
						Qvalue
16) Acetone	2.391	43	1550	1.974	ug/l	93
44) Trichloroethene	7.128	130	934	0.565	ug/l	69
64) Tetrachloroethene	9.280	164	1134	0.675	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092121\
Data File : VX024412.D
Acq On : 22 Sep 2021 02:38
Operator : JC/MD
Sample : M3833-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW202D1-20210919

Quant Time: Sep 22 04:56:19 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 21 10:52:48 2021
Response via : Initial Calibration

