

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092121\
 Data File : VX024414.D
 Acq On : 22 Sep 2021 03:25
 Operator : JC/MD
 Sample : M3833-14
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE122D3-20210920

Quant Time: Sep 22 04:56:42 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 21 10:52:48 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	127030	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	221938	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	234748	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	88197	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	97532	53.843	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	107.680%
35) Dibromofluoromethane	5.397	113	75518	50.737	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.480%
50) Toluene-d8	8.653	98	276383	50.972	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.940%
62) 4-Bromofluorobenzene	11.085	95	105777	50.128	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.260%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	1767	2.288	ug/l	90
44) Trichloroethene	7.141	130	2804	1.705	ug/l	77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092121\
 Data File : VX024414.D
 Acq On : 22 Sep 2021 03:25
 Operator : JC/MD
 Sample : M3833-14
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE122D3-20210920

Quant Time: Sep 22 04:56:42 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 21 10:52:48 2021
 Response via : Initial Calibration

