

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092121\
 Data File : VX024416.D
 Acq On : 22 Sep 2021 04:12
 Operator : JC/MD
 Sample : M3833-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE103D1-20210920

Quant Time: Sep 22 04:57:07 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 21 10:52:48 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	140931	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	262579	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	252824	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	103029	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	100289	49.904	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	99.800%
35) Dibromofluoromethane	5.397	113	78808	44.753	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.500%
50) Toluene-d8	8.653	98	290178	45.233	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	90.460%#
62) 4-Bromofluorobenzene	11.085	95	103354	41.399	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	82.800%#
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.173	85	468	0.345	ug/l	99
7) Trichlorofluoromethane	1.892	101	889	0.376	ug/l #	61
9) 1,1,2-Trichlorotrifluo...	2.331	101	13233	9.538	ug/l	98
12) 1,1-Dichloroethene	2.325	96	4604	3.543	ug/l	81
27) cis-1,2-Dichloroethene	4.501	96	4832	2.941	ug/l	92
30) Chloroform	5.099	83	1502	0.519	ug/l #	54
44) Trichloroethene	7.135	130	1408462	723.976	ug/l	98
55) 1,1,2-Trichloroethane	9.159	97	801	0.414	ug/l #	79
64) Tetrachloroethene	9.275	164	13949	7.122	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092121\
Data File : VX024416.D
Acq On : 22 Sep 2021 04:12
Operator : JC/MD
Sample : M3833-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE103D1-20210920

Quant Time: Sep 22 04:57:07 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 21 10:52:48 2021
Response via : Initial Calibration

