

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092121\
 Data File : VX024418.D
 Acq On : 22 Sep 2021 04:59
 Operator : JC/MD
 Sample : M3833-10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE103D3-20210920

Quant Time: Sep 22 06:57:46 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 21 10:52:48 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	132831	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.775	114	248003	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	259893	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	110722	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	99701	52.637	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 105.280%	
35) Dibromofluoromethane	5.397	113	77155	46.389	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 92.780%	
50) Toluene-d8	8.653	98	311462	51.405	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 102.800%	
62) 4-Bromofluorobenzene	11.085	95	122802	52.080	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 104.160%	
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.337	101	8997	6.880	ug/l	97
12) 1,1-Dichloroethene	2.319	96	764	0.624	ug/l	100
16) Acetone	2.386	43	2644	3.274	ug/l #	83
27) cis-1,2-Dichloroethene	4.495	96	1690	1.091	ug/l	94
30) Chloroform	5.111	83	1558	0.571	ug/l	96
44) Trichloroethene	7.135	130	945447	514.540	ug/l	100
64) Tetrachloroethene	9.281	164	2156	1.071	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092121\
Data File : VX024418.D
Acq On : 22 Sep 2021 04:59
Operator : JC/MD
Sample : M3833-10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE103D3-20210920

Quant Time: Sep 22 06:57:46 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 21 10:52:48 2021
Response via : Initial Calibration

