

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092121\
 Data File : VX024423.D
 Acq On : 22 Sep 2021 06:57
 Operator : JC/MD
 Sample : M3833-11 50X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP01-20210920

Quant Time: Sep 22 08:35:24 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 21 10:52:48 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	144431	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	253598	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	228433	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	105290	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	107640	52.264	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 104.520%	
35) Dibromofluoromethane	5.397	113	85444	50.239	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.480%	
50) Toluene-d8	8.653	98	269314	43.468	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 86.940%#	
62) 4-Bromofluorobenzene	11.085	95	115587	47.938	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 95.880%	
Target Compounds						
20) Methylene Chloride	2.800	84	1431	0.849	ug/l #	94
44) Trichloroethene	7.135	130	14663	7.804	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092121\
Data File : VX024423.D
Acq On : 22 Sep 2021 06:57
Operator : JC/MD
Sample : M3833-11 50X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 49 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP01-20210920

Quant Time: Sep 22 08:35:24 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 21 10:52:48 2021
Response via : Initial Calibration

