

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092121\
 Data File : VX024424.D
 Acq On : 22 Sep 2021 07:20
 Operator : JC/MD
 Sample : M3833-06 50X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP08-20210919

Quant Time: Sep 22 08:35:39 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 21 10:52:48 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	125837	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	220648	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	255791	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	108821	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.971	65	95470	53.205	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	106.400%
35) Dibromofluoromethane	5.397	113	73422	49.618	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.240%
50) Toluene-d8	8.653	98	278653	51.691	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	103.380%
62) 4-Bromofluorobenzene	11.085	95	120389	57.386	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	114.780%
Target Compounds						
						Qvalue
20) Methylene Chloride	2.794	84	1339	0.911	ug/l #	85
44) Trichloroethene	7.135	130	1058	0.647	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092121\
Data File : VX024424.D
Acq On : 22 Sep 2021 07:20
Operator : JC/MD
Sample : M3833-06 50X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 50 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP08-20210919

Quant Time: Sep 22 08:35:39 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 21 10:52:48 2021
Response via : Initial Calibration

