

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092122\
 Data File : VX031512.D
 Acq On : 21 Sep 2022 10:00
 Operator : JC/MD
 Sample : VX0921WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0921WBL01

Quant Time: Sep 22 01:10:55 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 20 09:45:29 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	52634	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	83167	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	86864	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	55930	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	51516	53.342	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	106.680%
35) Dibromofluoromethane	5.385	113	43328	50.410	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.820%
50) Toluene-d8	8.647	98	159017	47.991	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.980%
62) 4-Bromofluorobenzene	11.079	95	50486	44.982	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.960%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092122\
Data File : VX031512.D
Acq On : 21 Sep 2022 10:00
Operator : JC/MD
Sample : VX0921WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0921WBL01

Quant Time: Sep 22 01:10:55 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091922W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 20 09:45:29 2022
Response via : Initial Calibration

