

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092122\
 Data File : VX031516.D
 Acq On : 21 Sep 2022 11:59
 Operator : JC/MD
 Sample : N4583-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-5

Quant Time: Sep 22 01:11:57 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 20 09:45:29 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	77686	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	129678	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	114488	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	52570	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	88941	62.582	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 125.160%#	
35) Dibromofluoromethane	5.385	113	75312	56.279	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 112.560%	
50) Toluene-d8	8.647	98	292478	56.610	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 113.220%#	
62) 4-Bromofluorobenzene	11.079	95	89066	50.894	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 101.780%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092122\
Data File : VX031516.D
Acq On : 21 Sep 2022 11:59
Operator : JC/MD
Sample : N4583-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-5

Quant Time: Sep 22 01:11:57 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091922W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 20 09:45:29 2022
Response via : Initial Calibration

