

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092122\
 Data File : VX031521.D
 Acq On : 21 Sep 2022 13:55
 Operator : JC/MD
 Sample : N4678-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 5869-1

Quant Time: Sep 22 01:12:53 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 20 09:45:29 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	74239	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	124408	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	112223	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	51946	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	78022	57.358	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 114.720%	
35) Dibromofluoromethane	5.391	113	65114	50.647	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.300%	
50) Toluene-d8	8.653	98	252294	50.901	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 101.800%	
62) 4-Bromofluorobenzene	11.079	95	79730	47.489	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 94.980%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092122\
Data File : VX031521.D
Acq On : 21 Sep 2022 13:55
Operator : JC/MD
Sample : N4678-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
5869-1

Quant Time: Sep 22 01:12:53 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091922W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 20 09:45:29 2022
Response via : Initial Calibration

