

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092122\
 Data File : VX031533.D
 Acq On : 21 Sep 2022 18:37
 Operator : JC/MD
 Sample : N4614-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-3RRA

Quant Time: Sep 22 01:15:39 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 20 09:45:29 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	77572	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	128513	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	115770	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	54698	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	82222	57.858	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 115.720%	
35) Dibromofluoromethane	5.391	113	70729	53.295	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 106.580%	
50) Toluene-d8	8.653	98	272209	53.165	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 106.320%	
62) 4-Bromofluorobenzene	11.079	95	87036	50.184	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 100.360%	
Target Compounds						
					Qvalue	
4) Vinyl Chloride	1.368	62	6902	4.091	ug/l	95
12) 1,1-Dichloroethene	2.319	96	382	0.316	ug/l	80
19) Methyl tert-butyl Ether	3.124	73	1310	0.300	ug/l #	52
21) trans-1,2-Dichloroethene	3.087	96	1579	1.158	ug/l #	90
27) cis-1,2-Dichloroethene	4.489	96	106048	64.533	ug/l	80
44) Trichloroethene	7.129	130	67710	39.186	ug/l	89
64) Tetrachloroethene	9.275	164	454021	326.506	ug/l	98
67) Ethyl Benzene	10.195	91	1622	0.257	ug/l	100
78) n-propylbenzene	11.305	91	1216	0.259	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092122\
Data File : VX031533.D
Acq On : 21 Sep 2022 18:37
Operator : JC/MD
Sample : N4614-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-3RRA

Quant Time: Sep 22 01:15:39 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091922W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 20 09:45:29 2022
Response via : Initial Calibration

