

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092122\
 Data File : VX031547.D
 Acq On : 22 Sep 2022 04:09
 Operator : JC/MD
 Sample : N4642-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-1R

Quant Time: Sep 22 04:45:40 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 20 09:45:29 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	77027	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	125352	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	112557	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	50962	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	76141	53.884	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 107.760%	
35) Dibromofluoromethane	5.391	113	64856	50.058	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.120%	
50) Toluene-d8	8.653	98	250024	50.063	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 100.120%	
62) 4-Bromofluorobenzene	11.079	95	75238	44.476	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 88.960%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092122\
Data File : VX031547.D
Acq On : 22 Sep 2022 04:09
Operator : JC/MD
Sample : N4642-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-1R

Quant Time: Sep 22 04:45:40 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091922W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 20 09:45:29 2022
Response via : Initial Calibration

