

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092122\
 Data File : VX031548.D
 Acq On : 22 Sep 2022 04:33
 Operator : JC/MD
 Sample : N4642-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-2R

Quant Time: Sep 22 09:21:20 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 20 09:45:29 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	75965	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	123952	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	111365	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	49505	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	75613	54.266	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	108.540%
35) Dibromofluoromethane	5.391	113	64852	50.628	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.260%
50) Toluene-d8	8.653	98	246289	49.873	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.740%
62) 4-Bromofluorobenzene	11.079	95	76314	45.621	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.240%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092122\
Data File : VX031548.D
Acq On : 22 Sep 2022 04:33
Operator : JC/MD
Sample : N4642-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-2R

Quant Time: Sep 22 09:21:20 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091922W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 20 09:45:29 2022
Response via : Initial Calibration

