

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092122\
 Data File : VX031550.D
 Acq On : 22 Sep 2022 05:19
 Operator : JC/MD
 Sample : N4642-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-4R

Quant Time: Sep 22 09:21:44 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 20 09:45:29 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	69882	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	115988	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	104936	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	49371	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	68771	53.639	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	107.280%
35) Dibromofluoromethane	5.385	113	59194	49.366	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.740%
50) Toluene-d8	8.647	98	227131	49.151	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.300%
62) 4-Bromofluorobenzene	11.079	95	69895	44.653	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.300%
Target Compounds						
						Qvalue
7) Trichlorofluoromethane	1.892	101	910	0.264	ug/l	96
30) Chloroform	5.093	83	567	0.238	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092122\
Data File : VX031550.D
Acq On : 22 Sep 2022 05:19
Operator : JC/MD
Sample : N4642-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-4R

Quant Time: Sep 22 09:21:44 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091922W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 20 09:45:29 2022
Response via : Initial Calibration

