

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092122\
 Data File : VX031510.D
 Acq On : 21 Sep 2022 09:00
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 22 01:10:20 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 20 09:45:29 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	96	0.00
2 T	Dichlorodifluoromethane	0.719	0.768	-6.8	89	0.00
3 P	Chloromethane	0.954	0.852	10.7	89	0.00
4 C	Vinyl Chloride	1.087	1.026	5.6#	91	0.00
5 T	Bromomethane	1.548	1.907	-23.2	124	0.00
6 T	Chloroethane	1.174	1.234	-5.1	118	0.00
7 T	Trichlorofluoromethane	2.467	2.074	15.9	85	0.00
8 T	Diethyl Ether	0.850	0.694	18.4	85	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.805	0.755	6.2	95	0.00
10 T	Methyl Iodide	0.779	0.925	-18.7	102	0.00
11 T	Tert butyl alcohol	0.249	0.165	33.7#	77	0.00
12 CM	1,1-Dichloroethene	0.778	0.707	9.1#	92	0.00
13 T	Acrolein	0.114	0.084	26.3#	79	0.00
14 T	Allyl chloride	1.269	1.115	12.1	91	0.00
15 T	Acrylonitrile	0.483	0.415	14.1	85	0.00
16 T	Acetone	0.438	0.356	18.7	89	0.00
17 T	Carbon Disulfide	2.097	1.789	14.7	88	0.00
18 T	Methyl Acetate	1.370	1.137	17.0	87	0.00
19 T	Methyl tert-butyl Ether	2.815	2.542	9.7	91	0.00
20 T	Methylene Chloride	0.999	0.804	19.5	92	0.00
21 T	trans-1,2-Dichloroethene	0.879	0.803	8.6	91	0.00
22 T	Diisopropyl ether	2.748	2.510	8.7	91	0.00
23 T	Vinyl Acetate	2.350	2.179	7.3	90	0.00
24 P	1,1-Dichloroethane	1.553	1.403	9.7	91	0.00
25 T	2-Butanone	0.708	0.605	14.5	86	0.00
26 T	2,2-Dichloropropane	1.372	1.260	8.2	91	0.00
27 T	cis-1,2-Dichloroethene	1.059	0.934	11.8	91	0.00
28 T	Bromochloromethane	0.620	0.638	-2.9	102	0.00
29 T	Tetrahydrofuran	0.462	0.390	15.6	85	0.00
30 C	Chloroform	1.703	1.555	8.7#	89	0.00
31 T	Cyclohexane	1.360	1.245	8.5	91	0.00
32 T	1,1,1-Trichloroethane	1.547	1.384	10.5	89	0.00
33 S	1,2-Dichloroethane-d4	1.090	0.910	16.5	95	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	93	0.00
35 S	Dibromofluoromethane	0.592	0.522	11.8	94	0.00
36 T	1,1-Dichloropropene	0.795	0.793	0.3	93	0.00
37 T	Ethyl Acetate	0.948	0.821	13.4	86	0.00
38 T	Carbon Tetrachloride	0.881	0.845	4.1	91	0.00
39 T	Methylcyclohexane	0.978	0.946	3.3	93	0.00
40 TM	Benzene	2.371	2.250	5.1	90	0.00
41 T	Methacrylonitrile	0.472	0.414	12.3	86	-0.01
42 TM	1,2-Dichloroethane	0.874	0.836	4.3	91	0.00
43 T	Isopropyl Acetate	1.457	1.358	6.8	88	0.00
44 TM	Trichloroethene	0.672	0.637	5.2	93	0.00
45 C	1,2-Dichloropropane	0.596	0.561	5.9#	90	0.00
46 T	Dibromomethane	0.459	0.430	6.3	89	0.00
47 T	Bromodichloromethane	0.863	0.803	7.0	89	0.00
48 T	Methyl methacrylate	0.666	0.624	6.3	89	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092122\
 Data File : VX031510.D
 Acq On : 21 Sep 2022 09:00
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 22 01:10:20 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 20 09:45:29 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.015	0.013	13.3	79	0.00
50 S	Toluene-d8	1.992	1.995	-0.2	98	0.00
51 T	4-Methyl-2-Pentanone	0.937	0.863	7.9	85	0.00
52 CM	Toluene	1.546	1.512	2.2#	92	0.00
53 T	t-1,3-Dichloropropene	0.893	0.908	-1.7	89	0.00
54 T	cis-1,3-Dichloropropene	0.976	0.944	3.3	90	0.00
55 T	1,1,2-Trichloroethane	0.637	0.619	2.8	93	0.00
56 T	Ethyl methacrylate	0.906	0.922	-1.8	91	0.00
57 T	1,3-Dichloropropane	1.046	1.005	3.9	91	0.00
58 T	2-Chloroethyl Vinyl ether	0.350	0.351	-0.3	85	0.00
59 T	2-Hexanone	0.704	0.668	5.1	85	0.00
60 T	Dibromochloromethane	0.677	0.674	0.4	89	0.00
61 T	1,2-Dibromoethane	0.686	0.673	1.9	90	0.00
62 S	4-Bromofluorobenzene	0.675	0.721	-6.8	99	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	93	0.00
64 T	Tetrachloroethene	0.601	0.567	5.7	94	0.00
65 PM	Chlorobenzene	1.542	1.454	5.7	92	0.00
66 T	1,1,1,2-Tetrachloroethane	0.556	0.542	2.5	93	0.00
67 C	Ethyl Benzene	2.722	2.610	4.1#	92	0.00
68 T	m/p-Xylenes	1.094	1.058	3.3	93	0.00
69 T	o-Xylene	1.044	1.001	4.1	90	0.00
70 T	Styrene	1.719	1.711	0.5	92	0.00
71 P	Bromoform	0.445	0.431	3.1	89	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	87	0.00
73 T	Isopropylbenzene	3.734	3.527	5.5	93	0.00
74 T	N-amyl acetate	1.513	1.441	4.8	89	0.00
75 P	1,1,2,2-Tetrachloroethane	1.259	1.084	13.9	87	0.00
76 T	1,2,3-Trichloropropane	1.116	1.047	6.2	90	0.00
77 T	Bromobenzene	0.920	0.852	7.4	94	0.00
78 T	n-propylbenzene	4.296	4.198	2.3	93	0.00
79 T	2-Chlorotoluene	2.557	2.410	5.7	93	0.00
80 T	1,3,5-Trimethylbenzene	3.169	3.048	3.8	88	0.00
81 T	trans-1,4-Dichloro-2-butene	0.347	0.342	1.4	87	0.00
82 T	4-Chlorotoluene	3.001	2.888	3.8	86	0.00
83 T	tert-Butylbenzene	3.071	2.915	5.1	91	0.00
84 T	1,2,4-Trimethylbenzene	3.121	3.051	2.2	91	0.00
85 T	sec-Butylbenzene	3.887	3.799	2.3	92	0.00
86 T	p-Isopropyltoluene	3.282	3.276	0.2	90	0.00
87 T	1,3-Dichlorobenzene	1.764	1.677	4.9	92	0.00
88 T	1,4-Dichlorobenzene	1.840	1.799	2.2	91	0.00
89 T	n-Butylbenzene	2.796	2.851	-2.0	91	0.00
90 T	Hexachloroethane	0.564	0.552	2.1	91	0.00
91 T	1,2-Dichlorobenzene	1.776	1.661	6.5	90	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.257	0.240	6.6	86	0.00
93 T	1,2,4-Trichlorobenzene	0.979	0.979	0.0	94	0.00
94 T	Hexachlorobutadiene	0.391	0.381	2.6	101	0.00
95 T	Naphthalene	3.584	3.511	2.0	92	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092122\
Data File : VX031510.D
Acq On : 21 Sep 2022 09:00
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Sep 22 01:10:20 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091922W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 20 09:45:29 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.009	0.995	1.4	95	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6