



BFB

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092122\
Data File : VX031536.D
Acq On : 21 Sep 2022 21:57
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091922W.M
Title : SW846 8260
Last Update : Tue Sep 20 09:45:29 2022



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.9	3645	PASS
75	95	30	60	58.5	9745	PASS
95	95	100	100	100.0	16644	PASS
96	95	5	9	6.9	1155	PASS
173	174	0.00	2	1.2	185	PASS
174	95	50	100	89.0	14812	PASS
175	174	5	9	7.2	1066	PASS
176	174	95	101	98.9	14646	PASS
177	176	5	9	6.4	944	PASS