

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092123\
 Data File : VX037850.D
 Acq On : 21 Sep 2023 13:17
 Operator : JC/MD
 Sample : 04540-02 50X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EMBRIDGE-M-R-HEATER

Quant Time: Sep 22 04:42:57 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091523W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 16 04:55:44 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	232793	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	392582	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	378428	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	198981	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	143201	46.820	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.640%
35) Dibromofluoromethane	5.385	113	121248	46.158	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.320%
50) Toluene-d8	8.647	98	476771	48.849	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.700%
62) 4-Bromofluorobenzene	11.079	95	180911	51.009	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.020%
Target Compounds						
20) Methylene Chloride	2.788	84	1465	0.514	ug/l	# 84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092123\
Data File : VX037850.D
Acq On : 21 Sep 2023 13:17
Operator : JC/MD
Sample : 04540-02 50X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EMBRIDGE-M-R-HEATER

Quant Time: Sep 22 04:42:57 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091523W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 16 04:55:44 2023
Response via : Initial Calibration

