

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092123\
 Data File : VX037853.D
 Acq On : 21 Sep 2023 14:27
 Operator : JC/MD
 Sample : 04551-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GWOW-B6-L10-29-41-092023

Quant Time: Sep 22 04:44:10 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091523W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 16 04:55:44 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	223987	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	378526	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	372192	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	194210	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	140013	47.577	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.160%
35) Dibromofluoromethane	5.391	113	117903	46.552	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.100%
50) Toluene-d8	8.653	98	461836	49.076	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.160%
62) 4-Bromofluorobenzene	11.079	95	175099	51.203	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.400%
Target Compounds						
						Qvalue
16) Acetone	2.380	43	2994	2.115	ug/l	91
44) Trichloroethene	7.135	130	1929	0.646	ug/l	88
51) 4-Methyl-2-Pentanone	8.580	43	2952	0.759	ug/l	96
52) Toluene	8.720	92	182330	26.369	ug/l	93
64) Tetrachloroethene	9.281	164	1860	0.733	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092123\
Data File : VX037853.D
Acq On : 21 Sep 2023 14:27
Operator : JC/MD
Sample : 04551-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GWOW-B6-L10-29-41-092023

Quant Time: Sep 22 04:44:10 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091523W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 16 04:55:44 2023
Response via : Initial Calibration

