

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092123\
 Data File : VX037857.D
 Acq On : 21 Sep 2023 15:59
 Operator : JC/MD
 Sample : 04540-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TRANSCO-M-R-HEATER

Quant Time: Sep 22 04:45:56 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091523W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 16 04:55:44 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	212778	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	367947	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	354642	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	181175	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	136070	48.673	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.340%
35) Dibromofluoromethane	5.385	113	115812	47.041	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.080%
50) Toluene-d8	8.653	98	447919	48.965	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.940%
62) 4-Bromofluorobenzene	11.079	95	165390	49.755	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.500%
Target Compounds						
						Qvalue
20) Methylene Chloride	2.788	84	1693	0.656	ug/l	87
30) Chloroform	5.105	83	3099	0.667	ug/l	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092123\
Data File : VX037857.D
Acq On : 21 Sep 2023 15:59
Operator : JC/MD
Sample : 04540-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TRANSCO-M-R-HEATER

Quant Time: Sep 22 04:45:56 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091523W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 16 04:55:44 2023
Response via : Initial Calibration

