

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092123\
 Data File : VX037858.D
 Acq On : 21 Sep 2023 16:22
 Operator : JC/MD
 Sample : 04555-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-230920-01

Quant Time: Sep 22 05:58:25 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091523W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 16 04:55:44 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	218922	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	368705	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	352367	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	185290	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	135186	47.000	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.000%
35) Dibromofluoromethane	5.391	113	115397	46.776	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.560%
50) Toluene-d8	8.653	98	445322	48.581	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.160%
62) 4-Bromofluorobenzene	11.079	95	165206	49.597	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.200%
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	2.959	59	4420	7.159	ug/l	# 87
29) Tetrahydrofuran	5.019	42	7271	5.987	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092123\
Data File : VX037858.D
Acq On : 21 Sep 2023 16:22
Operator : JC/MD
Sample : 04555-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-230920-01

Quant Time: Sep 22 05:58:25 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091523W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 16 04:55:44 2023
Response via : Initial Calibration

