

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092123\
 Data File : VX037859.D
 Acq On : 21 Sep 2023 16:46
 Operator : JC/MD
 Sample : 04574-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FRAC-TANK-A2421

Quant Time: Sep 22 04:46:46 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091523W.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 16 04:55:44 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	212484	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	362806	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	346166	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	180150	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	134463	48.165	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.340%
35) Dibromofluoromethane	5.391	113	114419	47.133	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.260%
50) Toluene-d8	8.653	98	442550	49.064	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.120%
62) 4-Bromofluorobenzene	11.079	95	163328	49.831	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.660%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	15785	11.754	ug/l #	85
17) Carbon Disulfide	2.514	76	2759	0.531	ug/l	96
20) Methylene Chloride	2.794	84	1340	0.515	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092123\
Data File : VX037859.D
Acq On : 21 Sep 2023 16:46
Operator : JC/MD
Sample : 04574-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FRAC-TANK-A2421

Quant Time: Sep 22 04:46:46 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091523W.M
Quant Title : SW846 8260
QLast Update : Sat Sep 16 04:55:44 2023
Response via : Initial Calibration

