

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092221\
 Data File : VX024442.D
 Acq On : 22 Sep 2021 18:15
 Operator : JC/MD
 Sample : VX0922MBS01
 Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0922MBS01

Quant Time: Sep 23 06:40:10 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 21 10:52:48 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	181663	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	243252	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	243374	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	136040	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	130240	50.277	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 100.560%	
35) Dibromofluoromethane	5.397	113	104116	63.822	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 127.640%#	
50) Toluene-d8	8.653	98	320797	53.979	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 107.960%	
62) 4-Bromofluorobenzene	11.085	95	147743	63.881	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 127.760%#	
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.173	85	30265	17.331	ug/l	96
3) Chloromethane	1.294	50	32201	17.658	ug/l	99
4) Vinyl Chloride	1.374	62	32653	17.992	ug/l	99
5) Bromomethane	1.599	94	20257	18.749	ug/l	93
6) Chloroethane	1.685	64	18319	15.282	ug/l	90
7) Trichlorofluoromethane	1.886	101	49141	16.142	ug/l	95
8) Diethyl Ether	2.142	74	17086	16.220	ug/l	88
9) 1,1,2-Trichlorotrifluo...	2.331	101	34405	19.237	ug/l	97
10) Methyl Iodide	2.459	142	43465	17.567	ug/l	93
11) Tert butyl alcohol	2.977	59	63550	103.902	ug/l	98
12) 1,1-Dichloroethene	2.319	96	32084	19.155	ug/l	95
13) Acrolein	2.245	56	11392	66.724	ug/l	91
14) Allyl chloride	2.666	41	57679	18.948	ug/l	94
15) Acrylonitrile	3.075	53	111187	104.939	ug/l	97
16) Acetone	2.392	43	104004	94.159	ug/l	91
17) Carbon Disulfide	2.514	76	62989	15.006	ug/l	99
18) Methyl Acetate	2.715	43	73212	21.142	ug/l	92
19) Methyl tert-butyl Ether	3.123	73	124135	20.166	ug/l	99
20) Methylene Chloride	2.794	84	39349	18.552	ug/l	90
21) trans-1,2-Dichloroethene	3.099	96	34804	18.980	ug/l	97
22) Diisopropyl ether	3.770	45	124342	20.797	ug/l	90
23) Vinyl Acetate	3.733	43	485927	100.097	ug/l	94
24) 1,1-Dichloroethane	3.617	63	68824	19.617	ug/l	99
25) 2-Butanone	4.574	43	164551	101.145	ug/l	89
26) 2,2-Dichloropropane	4.489	77	56160	18.276	ug/l	99
27) cis-1,2-Dichloroethene	4.501	96	41455	19.575	ug/l	95
28) Bromochloromethane	4.910	49	23893	16.715	ug/l	98
29) Tetrahydrofuran	5.025	42	105361	103.470	ug/l	91
30) Chloroform	5.105	83	74732	20.031	ug/l	100
31) Cyclohexane	5.477	56	56109	18.517	ug/l	98
32) 1,1,1-Trichloroethane	5.397	97	64395	19.286	ug/l	99
36) 1,1-Dichloropropene	5.702	75	50471	22.567	ug/l	99
37) Ethyl Acetate	4.727	43	62008	23.795	ug/l	95
38) Carbon Tetrachloride	5.684	117	55133	21.902	ug/l	99
39) Methylcyclohexane	7.385	83	48046	17.994	ug/l	99
40) Benzene	6.050	78	136692	21.310	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092221\
 Data File : VX024442.D
 Acq On : 22 Sep 2021 18:15
 Operator : JC/MD
 Sample : VX0922MBS01
 Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0922MBS01

Quant Time: Sep 23 06:40:10 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 21 10:52:48 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.934	41	33741	24.279	ug/l	92
42) 1,2-Dichloroethane	6.098	62	58981	22.051	ug/l	94
43) Isopropyl Acetate	6.348	43	91247	21.996	ug/l	93
44) Trichloroethene	7.135	130	34412	19.094	ug/l	98
45) 1,2-Dichloropropane	7.440	63	34531	20.397	ug/l	98
46) Dibromomethane	7.586	93	24105	19.451	ug/l	99
47) Bromodichloromethane	7.830	83	44827	19.453	ug/l	96
48) Methyl methacrylate	7.696	41	41343	19.975	ug/l	84
49) 1,4-Dioxane	7.665	88	18110	417.871	ug/l #	95
51) 4-Methyl-2-Pentanone	8.580	43	284209	107.517	ug/l	89
52) Toluene	8.726	92	82765	19.687	ug/l	99
53) t-1,3-Dichloropropene	8.982	75	49482	18.935	ug/l	98
54) cis-1,3-Dichloropropene	8.372	75	50789	18.781	ug/l #	90
55) 1,1,2-Trichloroethane	9.153	97	36121	20.174	ug/l	96
56) Ethyl methacrylate	9.122	69	52744	18.529	ug/l #	79
57) 1,3-Dichloropropane	9.311	76	58681	20.125	ug/l	99
58) 2-Chloroethyl Vinyl ether	8.244	63	136710	105.510	ug/l	98
59) 2-Hexanone	9.433	43	225484	109.473	ug/l	86
60) Dibromochloromethane	9.525	129	34355	18.062	ug/l	99
61) 1,2-Dibromoethane	9.616	107	37571	19.693	ug/l	100
64) Tetrachloroethene	9.275	164	37904	20.104	ug/l	95
65) Chlorobenzene	10.086	112	97945	19.562	ug/l	100
66) 1,1,1,2-Tetrachloroethane	10.165	131	36612	19.730	ug/l	98
67) Ethyl Benzene	10.195	91	174985	19.978	ug/l	99
68) m/p-Xylenes	10.305	106	135925	40.535	ug/l	100
69) o-Xylene	10.646	106	73005	21.810	ug/l	99
70) Styrene	10.659	104	120802	22.247	ug/l	97
71) Bromoform	10.805	173	27784	18.664	ug/l #	100
73) Isopropylbenzene	10.963	105	194527	19.202	ug/l	99
74) N-amyl acetate	10.848	43	80701	18.923	ug/l	94
75) 1,1,2,2-Tetrachloroethane	11.213	83	68196	20.808	ug/l	99
76) 1,2,3-Trichloropropane	11.244	75	76308	24.394	ug/l	91
77) Bromobenzene	11.201	156	49464	19.428	ug/l	95
78) n-propylbenzene	11.305	91	219644	19.192	ug/l	99
79) 2-Chlorotoluene	11.366	91	136881	19.255	ug/l	98
80) 1,3,5-Trimethylbenzene	11.457	105	164270	19.237	ug/l	100
81) trans-1,4-Dichloro-2-b...	11.024	75	18660	17.544	ug/l	89
82) 4-Chlorotoluene	11.457	91	157300	19.448	ug/l	99
83) tert-Butylbenzene	11.719	119	161069	19.543	ug/l	96
84) 1,2,4-Trimethylbenzene	11.756	105	165444	19.696	ug/l	98
85) sec-Butylbenzene	11.896	105	204337	19.788	ug/l	99
86) p-Isopropyltoluene	12.012	119	169857	19.596	ug/l	98
87) 1,3-Dichlorobenzene	11.975	146	87936	19.282	ug/l	100
88) 1,4-Dichlorobenzene	12.043	146	87656	18.830	ug/l	96
89) n-Butylbenzene	12.335	91	141811	18.908	ug/l	99
90) Hexachloroethane	12.542	117	24707	17.691	ug/l	98
91) 1,2-Dichlorobenzene	12.335	146	88966	19.884	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.945	75	12761	16.972	ug/l	95
93) 1,2,4-Trichlorobenzene	13.591	180	50187	19.035	ug/l	97
94) Hexachlorobutadiene	13.731	225	24188	18.371	ug/l	99
95) Naphthalene	13.780	128	166780	18.000	ug/l	100
96) 1,2,3-Trichlorobenzene	13.963	180	50727	19.088	ug/l	98


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092221\
Data File : VX024442.D
Acq On : 22 Sep 2021 18:15
Operator : JC/MD
Sample : VX0922MBSD01
Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0922MBSD01

Quant Time: Sep 23 06:40:10 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 21 10:52:48 2021
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.562 min Scan# 734
Delta R.T. -0.000 min
Lab File: VX024442.D
Acq: 22 Sep 2021 18:15




Instrument :
MSVOA_X
ClientSampleId :
VX0922MBSD01


Tgt Ion:168 Resp: 181663
Ion Ratio Lower Upper
168 100
99 60.4 45.8 68.6



#2
Dichlorodifluoromethane
Concen: 17.331 ug/l
RT: 1.173 min Scan# 14
Delta R.T. -0.000 min
Lab File: VX024442.D
Acq: 22 Sep 2021 18:15

Tgt Ion: 85 Resp: 30265
Ion Ratio Lower Upper
85 100
87 30.8 16.4 49.2









#3
Chloromethane
Concen: 17.658 ug/l
RT: 1.294 min Scan# 34
Delta R.T. -0.001 min
Lab File: VX024442.D
Acq: 22 Sep 2021 18:15




Instrument :
MSVOA_X
ClientSampleId :
VX0922MBSD01


Tgt Ion: 50 Resp: 32201
Ion Ratio Lower Upper
50 100
52 33.4 26.4 39.6



#4
Vinyl Chloride
Concen: 17.992 ug/l
RT: 1.374 min Scan# 47
Delta R.T. -0.006 min
Lab File: VX024442.D
Acq: 22 Sep 2021 18:15

Tgt Ion: 62 Resp: 32653
Ion Ratio Lower Upper
62 100
64 32.5 25.7 38.5









#5
Bromomethane
Concen: 18.749 ug/l
RT: 1.599 min Scan# 84
Delta R.T. 0.000 min
Lab File: VX024442.D
Acq: 22 Sep 2021 18:15




Instrument :
MSVOA_X
ClientSampleId :
VX0922MBSD01


Tgt Ion: 94 Resp: 20257
Ion Ratio Lower Upper
94 100
96 88.9 76.4 114.6



#6
Chloroethane
Concen: 15.282 ug/l
RT: 1.685 min Scan# 98
Delta R.T. 0.006 min
Lab File: VX024442.D
Acq: 22 Sep 2021 18:15

Tgt Ion: 64 Resp: 18319
Ion Ratio Lower Upper
64 100
66 37.3 25.4 38.0









#7
Trichlorofluoromethane
Concen: 16.142 ug/l
RT: 1.886 min Scan# 131
Delta R.T. -0.000 min
Lab File: VX024442.D
Acq: 22 Sep 2021 18:15




Instrument :
MSVOA_X
ClientSampleId :
VX0922MBSD01


Tgt Ion:101 Resp: 49141
Ion Ratio Lower Upper
101 100
103 62.7 53.2 79.8



#8
Diethyl Ether
Concen: 16.220 ug/l
RT: 2.142 min Scan# 173
Delta R.T. -0.000 min
Lab File: VX024442.D
Acq: 22 Sep 2021 18:15

Tgt Ion: 74 Resp: 17086
Ion Ratio Lower Upper
74 100
45 99.9 44.5 133.5









#9
1,1,2-Trichlorotrifluoroethane
Concen: 19.237 ug/l
RT: 2.331 min Scan# 204
Delta R.T. -0.000 min
Lab File: VX024442.D
Acq: 22 Sep 2021 18:15




Instrument :
MSVOA_X
ClientSampleId :
VX0922MBSD01


Tgt Ion:101 Resp: 34405
Ion Ratio Lower Upper
101 100
85 44.8 34.9 52.3
151 80.0 61.4 92.2



#10
Methyl Iodide
Concen: 17.567 ug/l
RT: 2.459 min Scan# 225
Delta R.T. 0.006 min
Lab File: VX024442.D
Acq: 22 Sep 2021 18:15

Tgt Ion:142 Resp: 43465
Ion Ratio Lower Upper
142 100
127 47.2 33.3 49.9
141 14.4 11.4 17.0









#11
Tert butyl alcohol
Concen: 103.902 ug/l
RT: 2.977 min Scan# 310
Delta R.T. -0.006 min
Lab File: VX024442.D
Acq: 22 Sep 2021 18:15




Instrument :
MSVOA_X
ClientSampleId :
VX0922MBSD01


Tgt Ion: 59 Resp: 63550
Ion Ratio Lower Upper
59 100
57 9.1 7.8 11.8



#12
1,1-Dichloroethene
Concen: 19.155 ug/l
RT: 2.319 min Scan# 202
Delta R.T. -0.000 min
Lab File: VX024442.D
Acq: 22 Sep 2021 18:15

Tgt Ion: 96 Resp: 32084
Ion Ratio Lower Upper
96 100
61 164.8 124.6 187.0
98 63.8 50.6 76.0









#13
Acrolein
Concen: 66.724 ug/l
RT: 2.245 min Scan# 190
Delta R.T. -0.001 min
Lab File: VX024442.D
Acq: 22 Sep 2021 18:15




Instrument :
MSVOA_X
ClientSampleId :
VX0922MBSD01


Tgt Ion: 56 Resp: 11392
Ion Ratio Lower Upper
56 100
55 78.1 56.4 84.6



#14
Allyl chloride
Concen: 18.948 ug/l
RT: 2.666 min Scan# 259
Delta R.T. 0.000 min
Lab File: VX024442.D
Acq: 22 Sep 2021 18:15

Tgt Ion: 41 Resp: 57679
Ion Ratio Lower Upper
41 100
39 66.3 51.6 77.4
76 30.8 30.5 45.7









#15
Acrylonitrile
Concen: 104.939 ug/l
RT: 3.075 min Scan# 326
Delta R.T. -0.000 min
Lab File: VX024442.D
Acq: 22 Sep 2021 18:15




Instrument :
MSVOA_X
ClientSampleId :
VX0922MBSD01


Tgt Ion: 53 Resp: 111187
Ion Ratio Lower Upper
53 100
52 84.1 65.3 97.9
51 36.5 28.2 42.2



#16
Acetone
Concen: 94.159 ug/l
RT: 2.392 min Scan# 214
Delta R.T. -0.000 min
Lab File: VX024442.D
Acq: 22 Sep 2021 18:15

Tgt Ion: 43 Resp: 104004
Ion Ratio Lower Upper
43 100
58 29.7 27.8 41.8









#17
Carbon Disulfide
Concen: 15.006 ug/l
RT: 2.514 min Scan# 234
Delta R.T. -0.000 min
Lab File: VX024442.D
Acq: 22 Sep 2021 18:15




Instrument :
MSVOA_X
ClientSampleId :
VX0922MBSD01


Tgt Ion: 76 Resp: 62989
Ion Ratio Lower Upper
76 100
78 9.3 7.2 10.8



#18
Methyl Acetate
Concen: 21.142 ug/l
RT: 2.715 min Scan# 267
Delta R.T. -0.000 min
Lab File: VX024442.D
Acq: 22 Sep 2021 18:15

Tgt Ion: 43 Resp: 73212
Ion Ratio Lower Upper
43 100
74 23.3 21.9 32.9









#19
Methyl tert-butyl Ether
Concen: 20.166 ug/l
RT: 3.123 min Scan# 334
Delta R.T. 0.000 min
Lab File: VX024442.D
Acq: 22 Sep 2021 18:15




Instrument :
MSVOA_X
ClientSampleId :
VX0922MBSD01


Tgt Ion: 73 Resp: 124135
Ion Ratio Lower Upper
73 100
57 22.0 17.1 25.7



#20
Methylene Chloride
Concen: 18.552 ug/l
RT: 2.794 min Scan# 280
Delta R.T. 0.000 min
Lab File: VX024442.D
Acq: 22 Sep 2021 18:15

Tgt Ion: 84 Resp: 39349
Ion Ratio Lower Upper
84 100
49 126.4 88.6 132.8
51 35.0 27.0 40.6
86 67.2 50.3 75.5









#21
trans-1,2-Dichloroethene
Concen: 18.980 ug/l
RT: 3.099 min Scan# 330
Delta R.T. 0.000 min
Lab File: VX024442.D
Acq: 22 Sep 2021 18:15




Instrument :
MSVOA_X
ClientSampleId :
VX0922MBSD01


Tgt Ion: 96 Resp: 34804
Ion Ratio Lower Upper
96 100
61 137.5 107.7 161.5
98 66.6 50.4 75.6



#22
Diisopropyl ether
Concen: 20.797 ug/l
RT: 3.770 min Scan# 440
Delta R.T. -0.006 min
Lab File: VX024442.D
Acq: 22 Sep 2021 18:15

Tgt Ion: 45 Resp: 124342
Ion Ratio Lower Upper
45 100
43 83.5 58.0 87.0
87 28.3 25.4 38.0
59 12.3 10.2 15.2









#23
 Vinyl Acetate
 Concen: 100.097 ug/l
 RT: 3.733 min Scan# 434
 Delta R.T. 0.000 min
 Lab File: VX024442.D
 Acq: 22 Sep 2021 18:15




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0922MBSD01


Tgt Ion: 43 Resp: 485927
 Ion Ratio Lower Upper
 43 100
 86 10.2 9.9 14.9



#24
 1,1-Dichloroethane
 Concen: 19.617 ug/l
 RT: 3.617 min Scan# 415
 Delta R.T. 0.000 min
 Lab File: VX024442.D
 Acq: 22 Sep 2021 18:15

Tgt Ion: 63 Resp: 68824
 Ion Ratio Lower Upper
 63 100
 98 7.2 3.8 11.2
 100 4.3 2.5 7.5









#25
2-Butanone
Concen: 101.145 ug/l
RT: 4.574 min Scan# 572
Delta R.T. 0.006 min
Lab File: VX024442.D
Acq: 22 Sep 2021 18:15




Instrument :
MSVOA_X
ClientSampleId :
VX0922MBSD01


Tgt Ion: 43 Resp: 164551
Ion Ratio Lower Upper
43 100
72 23.7 23.7 35.5



#26
2,2-Dichloropropane
Concen: 18.276 ug/l
RT: 4.489 min Scan# 558
Delta R.T. 0.000 min
Lab File: VX024442.D
Acq: 22 Sep 2021 18:15

Tgt Ion: 77 Resp: 56160
Ion Ratio Lower Upper
77 100
97 22.8 11.7 35.1





#27

cis-1,2-Dichloroethene

Concen: 19.575 ug/l

RT: 4.501 min Scan# 560





Delta R.T. 0.006 min

Lab File: VX024442.D

Acq: 22 Sep 2021 18:15

Tgt Ion: 96 Resp: 41455

Ion	Ratio	Lower	Upper
96	100		
61	151.3	0.0	285.6
98	66.6	0.0	129.6



#28

Bromochloromethane

Concen: 16.715 ug/l

RT: 4.910 min Scan# 627




Delta R.T. -0.000 min


Lab File: VX024442.D

Acq: 22 Sep 2021 18:15

Tgt Ion: 49 Resp: 23893

Ion	Ratio	Lower	Upper
49	100		
129	2.1	0.0	4.0
130	83.1	68.2	102.2





#29

Tetrahydrofuran

Concen: 103.470 ug/l

RT: 5.025 min Scan# 646

Delta R.T. 0.006 min

Lab File: VX024442.D

Acq: 22 Sep 2021 18:15

Instrument :

MSVOA_X

ClientSampleId :

VX0922MBSD01





Tgt Ion: 42 Resp: 105361

Ion Ratio Lower Upper

42 100

72 44.8 41.7 62.5

71 42.5 38.4 57.6



#30

Chloroform

Concen: 20.031 ug/l

RT: 5.105 min Scan# 659

Delta R.T. -0.000 min

Lab File: VX024442.D




Acq: 22 Sep 2021 18:15


Tgt Ion: 83 Resp: 74732

Ion Ratio Lower Upper

83 100

85 64.9 52.0 78.0









#31
Cyclohexane
Concen: 18.517 ug/l
RT: 5.477 min Scan# 720
Delta R.T. -0.000 min
Lab File: VX024442.D
Acq: 22 Sep 2021 18:15




Instrument :
MSVOA_X
ClientSampleId :
VX0922MBSD01


Tgt Ion:	56	Resp:	56109
Ion	Ratio	Lower	Upper
56	100		
69	31.9	25.3	37.9
84	86.8	71.4	107.0



#32
1,1,1-Trichloroethane
Concen: 19.286 ug/l
RT: 5.397 min Scan# 707
Delta R.T. 0.000 min
Lab File: VX024442.D
Acq: 22 Sep 2021 18:15

Tgt Ion:	97	Resp:	64395
Ion	Ratio	Lower	Upper
97	100		
99	64.9	51.5	77.3
61	43.8	34.2	51.4





#33

1,2-Dichloroethane-d4

Concen: 50.277 ug/l

RT: 5.964 min Scan# 800

Delta R.T. -0.007 min

Lab File: VX024442.D





Acq: 22 Sep 2021 18:15

Tgt Ion: 65 Resp: 130240

Ion Ratio Lower Upper

65 100

67 50.8 0.0 105.4



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.769 min Scan# 932

Delta R.T. 0.000 min

Lab File: VX024442.D

Acq: 22 Sep 2021 18:15




Tgt Ion: 114 Resp: 243252


Ion Ratio Lower Upper

114 100

63 21.8 0.0 40.8

88 15.1 0.0 32.8









#35
Dibromofluoromethane
Concen: 63.822 ug/l
RT: 5.397 min Scan# 707
Delta R.T. 0.000 min
Lab File: VX024442.D
Acq: 22 Sep 2021 18:15




Instrument :
MSVOA_X
ClientSampleId :
VX0922MBSD01


Tgt Ion:	113	Resp:	104116
Ion	Ratio	Lower	Upper
113	100		
111	102.6	81.9	122.9
192	19.4	16.7	25.1



#36
1,1-Dichloropropene
Concen: 22.567 ug/l
RT: 5.702 min Scan# 757
Delta R.T. 0.000 min
Lab File: VX024442.D
Acq: 22 Sep 2021 18:15

Tgt Ion:	75	Resp:	50471
Ion	Ratio	Lower	Upper
75	100		
110	36.6	18.1	54.4
77	30.9	25.3	37.9





#37

Ethyl Acetate

Concen: 23.795 ug/l

RT: 4.727 min Scan# 597

Delta R.T. -0.006 min

Lab File: VX024442.D

Acq: 22 Sep 2021 18:15





Instrument :

MSVOA_X

ClientSampleId :

VX0922MBSD01

Tgt Ion:	43	Resp:	62008
Ion	Ratio	Lower	Upper
43	100		
61	14.0	12.8	19.2
70	10.7	10.5	15.7



#38

Carbon Tetrachloride

Concen: 21.902 ug/l




RT: 5.684 min Scan# 754


Delta R.T. -0.000 min

Lab File: VX024442.D

Acq: 22 Sep 2021 18:15

Tgt Ion:	117	Resp:	55133
Ion	Ratio	Lower	Upper
117	100		
119	95.7	76.2	114.2
121	32.7	24.8	37.2





#39

Methylcyclohexane

Concen: 17.994 ug/l

RT: 7.385 min Scan# 1033

Delta R.T. -0.000 min

Lab File: VX024442.D

Acq: 22 Sep 2021 18:15





Tgt Ion: 83 Resp: 48046

Ion Ratio Lower Upper

83 100

55 75.3 58.9 88.3

98 47.4 38.1 57.1



#40

Benzene

Concen: 21.310 ug/l

RT: 6.050 min Scan# 814

Delta R.T. -0.000 min

Lab File: VX024442.D




Acq: 22 Sep 2021 18:15


Tgt Ion: 78 Resp: 136692

Ion Ratio Lower Upper

78 100

77 23.9 18.9 28.3









#41
Methacrylonitrile
Concen: 24.279 ug/l
RT: 4.934 min Scan# 631
Delta R.T. 0.000 min
Lab File: VX024442.D
Acq: 22 Sep 2021 18:15




Instrument :
MSVOA_X
ClientSampleId :
VX0922MBSD01


Tgt Ion:	41	Resp:	33741
Ion	Ratio	Lower	Upper
41	100		
39	56.3	42.5	63.7
67	70.0	65.0	97.6
52	31.3	25.3	37.9



#42
1,2-Dichloroethane
Concen: 22.051 ug/l
RT: 6.098 min Scan# 822
Delta R.T. -0.001 min
Lab File: VX024442.D
Acq: 22 Sep 2021 18:15

Tgt Ion:	62	Resp:	58981
Ion	Ratio	Lower	Upper
62	100		
98	7.7	0.0	20.2









#43
Isopropyl Acetate
Concen: 21.996 ug/l
RT: 6.348 min Scan# 863
Delta R.T. -0.001 min
Lab File: VX024442.D
Acq: 22 Sep 2021 18:15




Instrument :
MSVOA_X
ClientSampleId :
VX0922MBSD01


Tgt Ion: 43	Resp: 91247
Ion Ratio	Lower Upper
43 100	
61 19.0	18.2 27.4
87 13.2	12.5 18.7



#44
Trichloroethene
Concen: 19.094 ug/l
RT: 7.135 min Scan# 992
Delta R.T. -0.000 min
Lab File: VX024442.D
Acq: 22 Sep 2021 18:15

Tgt Ion: 130	Resp: 34412
Ion Ratio	Lower Upper
130 100	
95 96.9	0.0 198.0





#45

1,2-Dichloropropane

Concen: 20.397 ug/l

RT: 7.440 min Scan# 1042

Delta R.T. -0.000 min

Lab File: VX024442.D





Acq: 22 Sep 2021 18:15

Tgt Ion: 63 Resp: 34531

Ion Ratio Lower Upper

63 100

65 33.2 25.7 38.5



#46

Dibromomethane

Concen: 19.451 ug/l

RT: 7.586 min Scan# 1066

Delta R.T. 0.000 min

Lab File: VX024442.D

Acq: 22 Sep 2021 18:15




Tgt Ion: 93 Resp: 24105


Ion Ratio Lower Upper

93 100

95 82.8 66.8 100.2

174 101.9 83.1 124.7





#47

Bromodichloromethane

Concen: 19.453 ug/l

RT: 7.830 min Scan# 1106





Delta R.T. -0.000 min

Lab File: VX024442.D

Acq: 22 Sep 2021 18:15

Tgt Ion:	83	Resp:	44827
Ion	Ratio	Lower	Upper
83	100		
85	66.9	51.4	77.0
127	9.9	6.8	10.2

Instrument :
MSVOA_X
ClientSampleId :
VX0922MBSD01



#48

Methyl methacrylate

Concen: 19.975 ug/l




RT: 7.696 min Scan# 1084


Delta R.T. -0.006 min

Lab File: VX024442.D

Acq: 22 Sep 2021 18:15

Tgt Ion:	41	Resp:	41343
Ion	Ratio	Lower	Upper
41	100		
69	77.3	76.0	114.0
39	64.4	44.6	67.0









#49
1,4-Dioxane
Concen: 417.871 ug/l
RT: 7.665 min Scan# 1079
Delta R.T. 0.000 min
Lab File: VX024442.D
Acq: 22 Sep 2021 18:15




Instrument :
MSVOA_X
ClientSampleId :
VX0922MBSD01


Tgt Ion: 88 Resp: 18110
Ion Ratio Lower Upper
88 100
43 35.9 23.2 34.8#
58 69.7 55.0 82.4



#50
Toluene-d8
Concen: 53.979 ug/l
RT: 8.653 min Scan# 1241
Delta R.T. -0.000 min
Lab File: VX024442.D
Acq: 22 Sep 2021 18:15

Tgt Ion: 98 Resp: 320797
Ion Ratio Lower Upper
98 100
100 64.2 51.9 77.9









#51
4-Methyl-2-Pentanone
Concen: 107.517 ug/l
RT: 8.580 min Scan# 1229
Delta R.T. -0.000 min
Lab File: VX024442.D
Acq: 22 Sep 2021 18:15




Instrument :
MSVOA_X
ClientSampleId :
VX0922MBSD01


Tgt Ion: 43 Resp: 284209
Ion Ratio Lower Upper
43 100
58 37.2 35.4 53.0



#52
Toluene
Concen: 19.687 ug/l
RT: 8.726 min Scan# 1253
Delta R.T. 0.000 min
Lab File: VX024442.D
Acq: 22 Sep 2021 18:15

Tgt Ion: 92 Resp: 82765
Ion Ratio Lower Upper
92 100
91 173.4 137.1 205.7









#53
t-1,3-Dichloropropene
Concen: 18.935 ug/l
RT: 8.982 min Scan# 1295
Delta R.T. 0.000 min
Lab File: VX024442.D
Acq: 22 Sep 2021 18:15




Instrument :
MSVOA_X
ClientSampleId :
VX0922MBSD01


Tgt Ion: 75 Resp: 49482
Ion Ratio Lower Upper
75 100
77 30.3 24.9 37.3



#54
cis-1,3-Dichloropropene
Concen: 18.781 ug/l
RT: 8.372 min Scan# 1195
Delta R.T. -0.001 min
Lab File: VX024442.D
Acq: 22 Sep 2021 18:15

Tgt Ion: 75 Resp: 50789
Ion Ratio Lower Upper
75 100
77 30.8 25.7 38.5
39 53.7 35.1 52.7#









#55
1,1,2-Trichloroethane
Concen: 20.174 ug/l
RT: 9.153 min Scan# 1323
Delta R.T. -0.006 min
Lab File: VX024442.D
Acq: 22 Sep 2021 18:15




Instrument :
MSVOA_X
ClientSampleId :
VX0922MBSD01


Tgt Ion: 97	Resp: 36121
Ion Ratio	Lower Upper
97 100	
83 79.6	69.3 103.9
85 56.9	44.3 66.5
99 64.6	50.3 75.5



#56
Ethyl methacrylate
Concen: 18.529 ug/l
RT: 9.122 min Scan# 1318
Delta R.T. 0.000 min
Lab File: VX024442.D
Acq: 22 Sep 2021 18:15

Tgt Ion: 69	Resp: 52744		
Ion Ratio	Lower	Upper	
69 100			
41 72.8	46.8	70.2#	
39 44.3	24.7	37.1#	









#57

1,3-Dichloropropane
Concen: 20.125 ug/l
RT: 9.311 min Scan# 1349
Delta R.T. 0.000 min
Lab File: VX024442.D
Acq: 22 Sep 2021 18:15

Instrument :
MSVOA_X
ClientSampleId :
VX0922MBSD01




Tgt Ion: 76 Resp: 58681
Ion Ratio Lower Upper
76 100
78 33.0 26.2 39.2




#58

2-Chloroethyl Vinyl ether
Concen: 105.510 ug/l
RT: 8.244 min Scan# 1174
Delta R.T. -0.007 min
Lab File: VX024442.D
Acq: 22 Sep 2021 18:15

Tgt Ion: 63 Resp: 136710
Ion Ratio Lower Upper
63 100
106 28.1 23.5 35.3









#59
2-Hexanone
Concen: 109.473 ug/l
RT: 9.433 min Scan# 1369
Delta R.T. 0.000 min
Lab File: VX024442.D
Acq: 22 Sep 2021 18:15




Instrument :
MSVOA_X
ClientSampleId :
VX0922MBSD01


Tgt Ion: 43 Resp: 225484
Ion Ratio Lower Upper
43 100
58 50.8 30.9 92.8



#60
Dibromochloromethane
Concen: 18.062 ug/l
RT: 9.525 min Scan# 1384
Delta R.T. -0.000 min
Lab File: VX024442.D
Acq: 22 Sep 2021 18:15

Tgt Ion: 129 Resp: 34355
Ion Ratio Lower Upper
129 100
127 78.0 38.4 115.2









#61
1,2-Dibromoethane
Concen: 19.693 ug/l
RT: 9.616 min Scan# 1399
Delta R.T. 0.000 min
Lab File: VX024442.D
Acq: 22 Sep 2021 18:15




Instrument :
MSVOA_X
ClientSampleId :
VX0922MBSD01


Tgt Ion:107 Resp: 37571
Ion Ratio Lower Upper
107 100
109 94.9 75.8 113.6



#62
4-Bromofluorobenzene
Concen: 63.881 ug/l
RT: 11.085 min Scan# 1640
Delta R.T. 0.000 min
Lab File: VX024442.D
Acq: 22 Sep 2021 18:15

Tgt Ion: 95 Resp: 147743
Ion Ratio Lower Upper
95 100
174 80.2 0.0 154.6
176 78.2 0.0 155.8









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. -0.006 min
Lab File: VX024442.D
Acq: 22 Sep 2021 18:15




Instrument :
MSVOA_X
ClientSampleId :
VX0922MBSD01


Tgt Ion:117 Resp: 243374
Ion Ratio Lower Upper
117 100
82 56.3 47.4 71.2
119 32.1 26.6 39.8



#64
Tetrachloroethene
Concen: 20.104 ug/l
RT: 9.275 min Scan# 1343
Delta R.T. -0.006 min
Lab File: VX024442.D
Acq: 22 Sep 2021 18:15

Tgt Ion:164 Resp: 37904
Ion Ratio Lower Upper
164 100
166 119.6 103.4 155.2
129 92.3 70.8 106.2
131 89.3 72.5 108.7









#65
Chlorobenzene
Concen: 19.562 ug/l
RT: 10.086 min Scan# 1476
Delta R.T. -0.000 min
Lab File: VX024442.D
Acq: 22 Sep 2021 18:15




Instrument :
MSVOA_X
ClientSampleId :
VX0922MBSD01


Tgt Ion:112 Resp: 97945
Ion Ratio Lower Upper
112 100
114 32.1 25.8 38.6



#66
1,1,1,2-Tetrachloroethane
Concen: 19.730 ug/l
RT: 10.165 min Scan# 1489
Delta R.T. -0.000 min
Lab File: VX024442.D
Acq: 22 Sep 2021 18:15

Tgt Ion:131 Resp: 36612
Ion Ratio Lower Upper
131 100
133 92.3 47.7 143.1
119 64.0 32.1 96.5









#67
Ethyl Benzene
Concen: 19.978 ug/l
RT: 10.195 min Scan# 1494
Delta R.T. 0.000 min
Lab File: VX024442.D
Acq: 22 Sep 2021 18:15




Instrument :
MSVOA_X
ClientSampleId :
VX0922MBSD01


Tgt Ion: 91 Resp: 174985
Ion Ratio Lower Upper
91 100
106 31.3 24.6 37.0



#68
m/p-Xylenes
Concen: 40.535 ug/l
RT: 10.305 min Scan# 1512
Delta R.T. 0.000 min
Lab File: VX024442.D
Acq: 22 Sep 2021 18:15

Tgt Ion: 106 Resp: 135925
Ion Ratio Lower Upper
106 100
91 204.2 162.8 244.2









#69
o-Xylene
Concen: 21.810 ug/l
RT: 10.646 min Scan# 1568
Delta R.T. -0.001 min
Lab File: VX024442.D
Acq: 22 Sep 2021 18:15




Instrument :
MSVOA_X
ClientSampleId :
VX0922MBSD01


Tgt Ion:106 Resp: 73005
Ion Ratio Lower Upper
106 100
91 214.9 108.1 324.2



#70
Styrene
Concen: 22.247 ug/l
RT: 10.659 min Scan# 1570
Delta R.T. -0.000 min
Lab File: VX024442.D
Acq: 22 Sep 2021 18:15

Tgt Ion:104 Resp: 120802
Ion Ratio Lower Upper
104 100
78 52.0 40.0 60.0
103 56.0 43.4 65.0





#71

Bromoform

Concen: 18.664 ug/l

RT: 10.805 min Scan# 1594





Delta R.T. -0.000 min

Lab File: VX024442.D

Acq: 22 Sep 2021 18:15

Tgt Ion:	173	Resp:	27784
Ion Ratio	Lower	Upper	
173	100		
175	47.9	23.9	71.8
254	0.0	0.1	0.1#

Instrument :
MSVOA_X
ClientSampleId :
VX0922MBSD01



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l




RT: 12.024 min Scan# 1794


Delta R.T. 0.000 min

Lab File: VX024442.D

Acq: 22 Sep 2021 18:15

Tgt Ion:	152	Resp:	136040
Ion Ratio	Lower	Upper	
152	100		
115	68.5	41.1	123.3
150	159.5	0.0	349.0









#73
Isopropylbenzene
Concen: 19.202 ug/l
RT: 10.963 min Scan# 1620
Delta R.T. -0.001 min
Lab File: VX024442.D
Acq: 22 Sep 2021 18:15




Instrument :
MSVOA_X
ClientSampleId :
VX0922MBSD01


Tgt Ion:105 Resp: 194527
Ion Ratio Lower Upper
105 100
120 26.9 13.2 39.6



#74
N-ethyl acetate
Concen: 18.923 ug/l
RT: 10.848 min Scan# 1601
Delta R.T. -0.000 min
Lab File: VX024442.D
Acq: 22 Sep 2021 18:15

Tgt Ion: 43 Resp: 80701
Ion Ratio Lower Upper
43 100
70 42.7 39.0 58.6
55 25.9 21.5 32.3
61 24.6 22.1 33.1









#75
 1,1,2,2-Tetrachloroethane
 Concen: 20.808 ug/l
 RT: 11.213 min Scan# 1661
 Delta R.T. -0.001 min
 Lab File: VX024442.D
 Acq: 22 Sep 2021 18:15




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0922MBSD01


Tgt Ion: 83 Resp: 68196
 Ion Ratio Lower Upper
 83 100
 131 10.8 5.1 15.2
 85 64.4 31.8 95.4



#76
 1,2,3-Trichloropropane
 Concen: 24.394 ug/l
 RT: 11.244 min Scan# 1666
 Delta R.T. -0.000 min
 Lab File: VX024442.D
 Acq: 22 Sep 2021 18:15

Tgt Ion: 75 Resp: 76308
 Ion Ratio Lower Upper
 75 100
 77 32.6 19.1 57.5









#77
Bromobenzene
Concen: 19.428 ug/l
RT: 11.201 min Scan# 1659
Delta R.T. 0.000 min
Lab File: VX024442.D
Acq: 22 Sep 2021 18:15




Instrument :
MSVOA_X
ClientSampleId :
VX0922MBSD01


Tgt Ion:156 Resp: 49464
Ion Ratio Lower Upper
156 100
77 156.3 82.8 248.6
158 97.2 49.1 147.4



#78
n-propylbenzene
Concen: 19.192 ug/l
RT: 11.305 min Scan# 1676
Delta R.T. -0.006 min
Lab File: VX024442.D
Acq: 22 Sep 2021 18:15

Tgt Ion: 91 Resp: 219644
Ion Ratio Lower Upper
91 100
120 23.5 11.5 34.5









#79
2-Chlorotoluene
Concen: 19.255 ug/l
RT: 11.366 min Scan# 1686
Delta R.T. -0.000 min
Lab File: VX024442.D
Acq: 22 Sep 2021 18:15




Instrument :
MSVOA_X
ClientSampleId :
VX0922MBSD01


Tgt Ion: 91 Resp: 136881
Ion Ratio Lower Upper
91 100
126 34.5 16.6 49.8



#80
1,3,5-Trimethylbenzene
Concen: 19.237 ug/l
RT: 11.457 min Scan# 1701
Delta R.T. 0.000 min
Lab File: VX024442.D
Acq: 22 Sep 2021 18:15

Tgt Ion: 105 Resp: 164270
Ion Ratio Lower Upper
105 100
120 47.4 23.8 71.5









#81
trans-1,4-Dichloro-2-butene
Concen: 17.544 ug/l
RT: 11.024 min Scan# 1630
Delta R.T. -0.001 min
Lab File: VX024442.D
Acq: 22 Sep 2021 18:15




Instrument :
MSVOA_X
ClientSampleId :
VX0922MBSD01


Tgt Ion: 75 Resp: 18660
Ion Ratio Lower Upper
75 100
53 108.0 74.5 111.7
89 45.3 37.9 56.9



#82
4-Chlorotoluene
Concen: 19.448 ug/l
RT: 11.457 min Scan# 1701
Delta R.T. 0.000 min
Lab File: VX024442.D
Acq: 22 Sep 2021 18:15

Tgt Ion: 91 Resp: 157300
Ion Ratio Lower Upper
91 100
126 29.2 14.9 44.5





#83

tert-Butylbenzene

Concen: 19.543 ug/l

RT: 11.719 min Scan# 1744

Delta R.T. -0.001 min

Lab File: VX024442.D

Acq: 22 Sep 2021 18:15

Tgt Ion:119 Resp: 161069

Ion Ratio Lower Upper

119 100

91 59.7 27.9 83.7





134 22.9 11.3 33.9

Instrument :

MSVOA_X

ClientSampleId :

VX0922MBSD01



#84

1,2,4-Trimethylbenzene

Concen: 19.696 ug/l

RT: 11.756 min Scan# 1750

Delta R.T. -0.000 min

Lab File: VX024442.D




Acq: 22 Sep 2021 18:15


Tgt Ion:105 Resp: 165444

Ion Ratio Lower Upper

105 100

120 43.9 22.4 67.3









#85
 sec-Butylbenzene
 Concen: 19.788 ug/l
 RT: 11.896 min Scan# 1773
 Delta R.T. 0.000 min
 Lab File: VX024442.D
 Acq: 22 Sep 2021 18:15




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0922MBSD01


Tgt Ion:105 Resp: 204337
 Ion Ratio Lower Upper
 105 100
 134 19.7 10.0 29.8



#86
 p-Isopropyltoluene
 Concen: 19.596 ug/l
 RT: 12.012 min Scan# 1792
 Delta R.T. 0.000 min
 Lab File: VX024442.D
 Acq: 22 Sep 2021 18:15

Tgt Ion:119 Resp: 169857
 Ion Ratio Lower Upper
 119 100
 134 26.2 12.9 38.7
 91 24.9 11.7 35.0









#87
1,3-Dichlorobenzene
Concen: 19.282 ug/l
RT: 11.975 min Scan# 1786
Delta R.T. -0.001 min
Lab File: VX024442.D
Acq: 22 Sep 2021 18:15




Instrument :
MSVOA_X
ClientSampleId :
VX0922MBSD01


Tgt Ion:146 Resp: 87936
Ion Ratio Lower Upper
146 100
111 40.1 20.2 60.5
148 64.1 31.9 95.7



#88
1,4-Dichlorobenzene
Concen: 18.830 ug/l
RT: 12.043 min Scan# 1797
Delta R.T. -0.000 min
Lab File: VX024442.D
Acq: 22 Sep 2021 18:15

Tgt Ion:146 Resp: 87656
Ion Ratio Lower Upper
146 100
111 42.2 19.6 58.8
148 66.1 31.6 95.0









#89
n-Butylbenzene
Concen: 18.908 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.000 min
Lab File: VX024442.D
Acq: 22 Sep 2021 18:15




Instrument :
MSVOA_X
ClientSampleId :
VX0922MBSD01


Tgt Ion: 91 Resp: 141811
Ion Ratio Lower Upper
91 100
92 53.6 27.4 82.0
134 26.0 13.2 39.5



#90
Hexachloroethane
Concen: 17.691 ug/l
RT: 12.542 min Scan# 1879
Delta R.T. -0.001 min
Lab File: VX024442.D
Acq: 22 Sep 2021 18:15

Tgt Ion: 117 Resp: 24707
Ion Ratio Lower Upper
117 100
201 79.0 40.4 121.2









#91
1,2-Dichlorobenzene
Concen: 19.884 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. -0.006 min
Lab File: VX024442.D
Acq: 22 Sep 2021 18:15




Instrument :
MSVOA_X
ClientSampleId :
VX0922MBSD01


Tgt Ion:146 Resp: 88966
Ion Ratio Lower Upper
146 100
111 41.8 20.8 62.2
148 63.2 31.9 95.5



#92
1,2-Dibromo-3-Chloropropane
Concen: 16.972 ug/l
RT: 12.945 min Scan# 1945
Delta R.T. -0.000 min
Lab File: VX024442.D
Acq: 22 Sep 2021 18:15

Tgt Ion: 75 Resp: 12761
Ion Ratio Lower Upper
75 100
155 85.8 45.3 135.9
157 110.6 57.7 173.1









#93
1,2,4-Trichlorobenzene
Concen: 19.035 ug/l
RT: 13.591 min Scan# 2051
Delta R.T. 0.000 min
Lab File: VX024442.D
Acq: 22 Sep 2021 18:15




Instrument :
MSVOA_X
ClientSampleId :
VX0922MBSD01


Tgt Ion:180	Resp:	50187
Ion Ratio	Lower	Upper
180	100	
182	93.2	47.7 143.1
145	32.2	15.0 45.0



#94
Hexachlorobutadiene
Concen: 18.371 ug/l
RT: 13.731 min Scan# 2074
Delta R.T. 0.006 min
Lab File: VX024442.D
Acq: 22 Sep 2021 18:15

Tgt Ion:225	Resp:	24188
Ion Ratio	Lower	Upper
225	100	
223	62.8	31.1 93.2
227	62.7	31.6 95.0









#95
Naphthalene
Concen: 18.000 ug/l
RT: 13.780 min Scan# 2082
Delta R.T. 0.000 min
Lab File: VX024442.D
Acq: 22 Sep 2021 18:15

Instrument :
MSVOA_X
ClientSampleId :
VX0922MBSD01

Tgt Ion:128 Resp: 166780
Ion Ratio Lower Upper
128 100
127 13.0 10.3 15.5
129 10.9 8.8 13.2



#96
1,2,3-Trichlorobenzene
Concen: 19.088 ug/l
RT: 13.963 min Scan# 2112
Delta R.T. -0.000 min
Lab File: VX024442.D
Acq: 22 Sep 2021 18:15

Tgt Ion:180 Resp: 50727
Ion Ratio Lower Upper
180 100
182 93.8 47.6 142.9
145 33.2 16.1 48.3

