

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092221\
 Data File : VX024431.D
 Acq On : 22 Sep 2021 13:50
 Operator : JC/MD
 Sample : M3835-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TRIP-BLANK

Quant Time: Sep 22 14:25:25 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 21 10:52:48 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	150449	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	254141	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	277213	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	104348	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	113598	52.951	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 105.900%	
35) Dibromofluoromethane	5.391	113	88044	51.658	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 103.320%	
50) Toluene-d8	8.653	98	309901	49.912	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 99.820%	
62) 4-Bromofluorobenzene	11.085	95	129708	53.680	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 107.360%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092221\
Data File : VX024431.D
Acq On : 22 Sep 2021 13:50
Operator : JC/MD
Sample : M3835-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TRIP-BLANK

Quant Time: Sep 22 14:25:25 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 21 10:52:48 2021
Response via : Initial Calibration

