

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092221\
 Data File : VX024432.D
 Acq On : 22 Sep 2021 14:20
 Operator : JC/MD
 Sample : M3786-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 N71924

Quant Time: Sep 22 15:18:02 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 21 10:52:48 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	145020	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	232088	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	263094	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	99982	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	111753	54.041	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 108.080%	
35) Dibromofluoromethane	5.397	113	83305	53.521	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 107.040%	
50) Toluene-d8	8.653	98	304437	53.691	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 107.380%	
62) 4-Bromofluorobenzene	11.085	95	122872	55.683	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 111.360%	
Target Compounds						
16) Acetone	2.392	43	5315	6.028	ug/l	Qvalue 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092221\
Data File : VX024432.D
Acq On : 22 Sep 2021 14:20
Operator : JC/MD
Sample : M3786-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
N71924

Quant Time: Sep 22 15:18:02 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 21 10:52:48 2021
Response via : Initial Calibration

