

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092221\
 Data File : VX024438.D
 Acq On : 22 Sep 2021 16:41
 Operator : JC/MD
 Sample : M3772-03ME 20X
 Misc : 5.97g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WASTE-PRE-CLASSME

Quant Time: Sep 22 17:04:47 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 21 10:52:48 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	148141	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	275357	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	244859	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	103182	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	109831	51.992	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	103.980%
35) Dibromofluoromethane	5.397	113	84589	45.806	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.620%
50) Toluene-d8	8.653	98	340663	50.639	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.280%
62) 4-Bromofluorobenzene	11.085	95	122449	46.771	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.540%
Target Compounds						Qvalue
64) Tetrachloroethene	9.275	164	1024	0.540	ug/l	# 84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092221\
Data File : VX024438.D
Acq On : 22 Sep 2021 16:41
Operator : JC/MD
Sample : M3772-03ME 20X
Misc : 5.97g/5mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WASTE-PRE-CLASSME

Quant Time: Sep 22 17:04:47 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 21 10:52:48 2021
Response via : Initial Calibration

