

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092221\
 Data File : VX024445.D
 Acq On : 22 Sep 2021 19:26
 Operator : JC/MD
 Sample : M3770-05DL 50X
 Misc : 6.76g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 NAPL-5-SIDEWALL-END-POINT-(10-

Quant Time: Sep 23 06:41:38 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 21 10:52:48 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	157664	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	272480	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	258617	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	115068	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	112986	50.255	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	100.520%
35) Dibromofluoromethane	5.397	113	90370	49.454	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.900%
50) Toluene-d8	8.653	98	341028	51.228	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	102.460%
62) 4-Bromofluorobenzene	11.085	95	125972	48.625	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.240%
Target Compounds						Qvalue
27) cis-1,2-Dichloroethene	4.495	96	31561	17.172	ug/l	93
44) Trichloroethene	7.135	130	17541	8.689	ug/l	94
64) Tetrachloroethene	9.275	164	80540	40.200	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092221\
Data File : VX024445.D
Acq On : 22 Sep 2021 19:26
Operator : JC/MD
Sample : M3770-05DL 50X
Misc : 6.76g/5mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
NAPL-5-SIDEWALL-END-POINT-(10-

Quant Time: Sep 23 06:41:38 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 21 10:52:48 2021
Response via : Initial Calibration

