

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092221\
 Data File : VX024447.D
 Acq On : 22 Sep 2021 20:13
 Operator : JC/MD
 Sample : M3833-17DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE126D2-20210920DL

Quant Time: Sep 23 06:42:17 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 21 10:52:48 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	141490	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	244722	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	272869	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	115740	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	104979	52.032	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	104.060%
35) Dibromofluoromethane	5.397	113	81412	49.605	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.200%
50) Toluene-d8	8.653	98	317663	53.131	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	106.260%
62) 4-Bromofluorobenzene	11.085	95	127675	54.872	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	109.740%
Target Compounds						
						Qvalue
44) Trichloroethene	7.135	130	85541	47.178	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092221\
Data File : VX024447.D
Acq On : 22 Sep 2021 20:13
Operator : JC/MD
Sample : M3833-17DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE126D2-20210920DL

Quant Time: Sep 23 06:42:17 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 21 10:52:48 2021
Response via : Initial Calibration

