



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092221\
Data File : VX024426.D
Acq On : 22 Sep 2021 09:15
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091421W.M
Title : SW846 8260
Last Update : Tue Sep 21 10:52:48 2021



AutoFind: Scans 1639, 1640, 1641; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.3	5096	PASS
75	95	30	60	54.5	14378	PASS
95	95	100	100	100.0	26405	PASS
96	95	5	9	7.0	1837	PASS
173	174	0.00	2	1.1	235	PASS
174	95	50	100	79.2	20912	PASS
175	174	5	9	7.8	1630	PASS
176	174	95	101	97.7	20427	PASS
177	176	5	9	7.0	1435	PASS