

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092222\
 Data File : VX031562.D
 Acq On : 22 Sep 2022 11:44
 Operator : JC/MD
 Sample : VX0922MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0922MBL01

Quant Time: Sep 22 17:59:41 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 20 09:45:29 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	69843	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.757	114	112201	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	103072	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	47179	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	77636	60.730	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 121.460%	
35) Dibromofluoromethane	5.385	113	67327	58.173	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 116.340%	
50) Toluene-d8	8.647	98	255380	57.129	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 114.260%#	
62) 4-Bromofluorobenzene	11.079	95	76145	50.288	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 100.580%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092222\
Data File : VX031562.D
Acq On : 22 Sep 2022 11:44
Operator : JC/MD
Sample : VX0922MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0922MBL01

Quant Time: Sep 22 17:59:41 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091922W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 20 09:45:29 2022
Response via : Initial Calibration

