

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092222\
 Data File : VX031568.D
 Acq On : 22 Sep 2022 14:24
 Operator : JC/MD
 Sample : N4566-04DL 500X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-GRDL

Quant Time: Sep 23 01:48:59 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 20 09:45:29 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	70759	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	116231	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	105254	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	49462	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	71171	54.847	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 109.700%	
35) Dibromofluoromethane	5.391	113	61135	50.901	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.800%	
50) Toluene-d8	8.653	98	230683	49.815	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 99.640%	
62) 4-Bromofluorobenzene	11.079	95	73268	46.710	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 93.420%	
Target Compounds						
					Qvalue	
52) Toluene	8.720	92	92073	25.613	ug/l	99
67) Ethyl Benzene	10.195	91	13246	2.311	ug/l	91
68) m/p-Xylenes	10.299	106	39836	17.297	ug/l	96
69) o-Xylene	10.646	106	19120	8.698	ug/l	92
80) 1,3,5-Trimethylbenzene	11.451	105	2369	0.756	ug/l	96
84) 1,2,4-Trimethylbenzene	11.756	105	10153	3.289	ug/l	97
95) Naphthalene	13.780	128	5490	1.548	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092222\
Data File : VX031568.D
Acq On : 22 Sep 2022 14:24
Operator : JC/MD
Sample : N4566-04DL 500X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-GRDL

Quant Time: Sep 23 01:48:59 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091922W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 20 09:45:29 2022
Response via : Initial Calibration

