

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092222\
 Data File : VX031569.D
 Acq On : 22 Sep 2022 14:47
 Operator : JC/MD
 Sample : N4642-11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-10

Quant Time: Sep 23 01:49:12 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 20 09:45:29 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	68341	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	111890	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	101292	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	46875	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	69092	55.135	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	110.260%
35) Dibromofluoromethane	5.391	113	58917	50.958	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.920%
50) Toluene-d8	8.653	98	227920	51.128	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	102.260%
62) 4-Bromofluorobenzene	11.079	95	67760	44.874	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.740%
Target Compounds						
30) Chloroform	5.099	83	10121	4.347	ug/l	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092222\
 Data File : VX031569.D
 Acq On : 22 Sep 2022 14:47
 Operator : JC/MD
 Sample : N4642-11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-10

Quant Time: Sep 23 01:49:12 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 20 09:45:29 2022
 Response via : Initial Calibration

