

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092222\
 Data File : VX031571.D
 Acq On : 22 Sep 2022 15:34
 Operator : JC/MD
 Sample : N4642-13
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FB091322

Quant Time: Sep 23 01:49:30 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 20 09:45:29 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	71771	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	115830	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	106878	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	47758	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	71841	54.578	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	109.160%
35) Dibromofluoromethane	5.385	113	60872	50.857	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.720%
50) Toluene-d8	8.653	98	232898	50.468	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.940%
62) 4-Bromofluorobenzene	11.079	95	72281	46.240	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.480%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092222\
Data File : VX031571.D
Acq On : 22 Sep 2022 15:34
Operator : JC/MD
Sample : N4642-13
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FB091322

Quant Time: Sep 23 01:49:30 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091922W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 20 09:45:29 2022
Response via : Initial Calibration

