

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092222\
 Data File : VX031576.D
 Acq On : 22 Sep 2022 17:31
 Operator : JC/MD
 Sample : N4728-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OF-1

Quant Time: Sep 23 01:49:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 20 09:45:29 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	69681	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	112598	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	102389	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	46050	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	70406	55.102	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	110.200%
35) Dibromofluoromethane	5.391	113	59755	51.364	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.720%
50) Toluene-d8	8.653	98	228921	51.030	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	102.060%
62) 4-Bromofluorobenzene	11.079	95	70113	46.141	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.280%
Target Compounds						
						Qvalue
16) Acetone	2.392	43	10560	7.655	ug/l	96
25) 2-Butanone	4.574	43	2989	3.030	ug/l	97
51) 4-Methyl-2-Pentanone	8.580	43	5496	2.604	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092222\
Data File : VX031576.D
Acq On : 22 Sep 2022 17:31
Operator : JC/MD
Sample : N4728-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
OF-1

Quant Time: Sep 23 01:49:49 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091922W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 20 09:45:29 2022
Response via : Initial Calibration

