

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092222\
 Data File : VX031578.D
 Acq On : 22 Sep 2022 18:18
 Operator : JC/MD
 Sample : N4728-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OF-3

Quant Time: Sep 23 01:50:09 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 20 09:45:29 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	66325	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	107687	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	99980	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	45736	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	68720	56.532	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	113.060%
35) Dibromofluoromethane	5.385	113	59064	53.110	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	106.220%
50) Toluene-d8	8.653	98	225868	52.646	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	105.300%
62) 4-Bromofluorobenzene	11.079	95	68903	47.412	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.820%
Target Compounds						
						Qvalue
16) Acetone	2.392	43	16335	21.471	ug/l	94
25) 2-Butanone	4.568	43	4741	5.049	ug/l #	85
51) 4-Methyl-2-Pentanone	8.580	43	8443	4.182	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092222\
Data File : VX031578.D
Acq On : 22 Sep 2022 18:18
Operator : JC/MD
Sample : N4728-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
OF-3

Quant Time: Sep 23 01:50:09 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091922W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 20 09:45:29 2022
Response via : Initial Calibration

