

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092222\
 Data File : VX031584.D
 Acq On : 22 Sep 2022 20:38
 Operator : JC/MD
 Sample : N4751-10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FB-9-21-22

Quant Time: Sep 23 01:51:21 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 20 09:45:29 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	71555	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	120046	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	106950	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	49548	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	71631	54.583	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 109.160%	
35) Dibromofluoromethane	5.391	113	60840	49.019	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 98.040%	
50) Toluene-d8	8.647	98	234750	49.083	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 98.160%	
62) 4-Bromofluorobenzene	11.079	95	72881	44.987	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 89.980%	
Target Compounds						
					Qvalue	
67) Ethyl Benzene	10.195	91	1475	0.253	ug/l	97
68) m/p-Xylenes	10.305	106	1583	0.676	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	820	0.261	ug/l	96
84) 1,2,4-Trimethylbenzene	11.756	105	2560	0.828	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092222\
Data File : VX031584.D
Acq On : 22 Sep 2022 20:38
Operator : JC/MD
Sample : N4751-10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FB-9-21-22

Quant Time: Sep 23 01:51:21 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091922W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 20 09:45:29 2022
Response via : Initial Calibration

