

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092222\
 Data File : VX031594.D
 Acq On : 23 Sep 2022 01:47
 Operator : JC/MD
 Sample : N4614-01DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 H-1RDL

Quant Time: Sep 23 04:47:06 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 20 09:45:29 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	78238	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	129279	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	117457	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	56789	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	79393	55.345	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 110.680%	
35) Dibromofluoromethane	5.385	113	67093	50.214	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.420%	
50) Toluene-d8	8.653	98	260793	50.633	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 101.260%	
62) 4-Bromofluorobenzene	11.079	95	80974	46.413	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 92.820%	
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	2.995	59	24758	56.963	ug/l	97
19) Methyl tert-butyl Ether	3.111	73	4257	0.967	ug/l #	1
22) Diisopropyl ether	3.764	45	4040	0.939	ug/l #	81
31) Cyclohexane	5.471	56	2625	1.233	ug/l	98
40) Benzene	6.038	78	11723	1.912	ug/l	98
64) Tetrachloroethene	9.269	164	915	0.649	ug/l #	76
67) Ethyl Benzene	10.195	91	13705	2.143	ug/l	92
68) m/p-Xylenes	10.305	106	2182	0.849	ug/l	94
73) Isopropylbenzene	10.963	105	25631	6.044	ug/l	100
78) n-propylbenzene	11.305	91	54344	11.139	ug/l	99
84) 1,2,4-Trimethylbenzene	11.750	105	67542	19.056	ug/l	98
85) sec-Butylbenzene	11.890	105	4587	1.039	ug/l	93
89) n-Butylbenzene	12.335	91	4681	1.474	ug/l #	70
95) Naphthalene	13.774	128	42208	10.368	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092222\
Data File : VX031594.D
Acq On : 23 Sep 2022 01:47
Operator : JC/MD
Sample : N4614-01DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
H-1RDL

Quant Time: Sep 23 04:47:06 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091922W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 20 09:45:29 2022
Response via : Initial Calibration

