

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092222\
 Data File : VX031597.D
 Acq On : 23 Sep 2022 02:57
 Operator : JC/MD
 Sample : N4614-05DL 25X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-3RRBDL

Quant Time: Sep 23 04:47:46 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 20 09:45:29 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	74817	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	123053	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	112915	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	50699	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	73618	53.632	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	107.260%
35) Dibromofluoromethane	5.385	113	62447	49.085	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.160%
50) Toluene-d8	8.647	98	243458	49.659	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.320%
62) 4-Bromofluorobenzene	11.079	95	74795	45.040	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.080%
Target Compounds						
						Qvalue
4) Vinyl Chloride	1.374	62	602	0.370	ug/l	# 69
27) cis-1,2-Dichloroethene	4.495	96	7956	5.020	ug/l	79
44) Trichloroethene	7.135	130	4832	2.920	ug/l	72
64) Tetrachloroethene	9.275	164	28183	20.780	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092222\
Data File : VX031597.D
Acq On : 23 Sep 2022 02:57
Operator : JC/MD
Sample : N4614-05DL 25X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-3RRBDL

Quant Time: Sep 23 04:47:46 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091922W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 20 09:45:29 2022
Response via : Initial Calibration

