

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092222\
 Data File : VX031599.D
 Acq On : 23 Sep 2022 03:44
 Operator : JC/MD
 Sample : PB147821TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB147821TB

Quant Time: Sep 23 04:48:08 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 20 09:45:29 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	73246	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	121855	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	123012	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	55467	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	72988	54.328	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 108.660%			
35) Dibromofluoromethane	5.391	113	62614	49.709	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 99.420%			
50) Toluene-d8	8.653	98	246895	50.856	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 101.720%			
62) 4-Bromofluorobenzene	11.079	95	79434	48.304	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 96.600%			
Target Compounds						
20) Methylene Chloride	2.788	84	4129	1.157	ug/l	88
37) Ethyl Acetate	4.721	43	10191	4.410	ug/l	97
43) Isopropyl Acetate	6.348	43	43054	12.123	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092222\
Data File : VX031599.D
Acq On : 23 Sep 2022 03:44
Operator : JC/MD
Sample : PB147821TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB147821TB

Quant Time: Sep 23 04:48:08 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091922W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 20 09:45:29 2022
Response via : Initial Calibration

