

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092222\
 Data File : VX031600.D
 Acq On : 23 Sep 2022 04:07
 Operator : JC/MD
 Sample : N4731-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SB-1

Quant Time: Sep 23 04:48:18 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 20 09:45:29 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	72712	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	126190	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	126381	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	99072	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	75160	56.396	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	112.800%
35) Dibromofluoromethane	5.385	113	64183	49.197	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.400%
50) Toluene-d8	8.653	98	259564	51.628	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	103.260%
62) 4-Bromofluorobenzene	11.079	95	129970	76.320	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	152.640%#
Target Compounds						Qvalue
16) Acetone	2.386	43	26461	38.626	ug/l	99
20) Methylene Chloride	2.788	84	8760	5.261	ug/l #	94
31) Cyclohexane	5.458	56	2181	1.103	ug/l	94
37) Ethyl Acetate	4.721	43	10348	4.324	ug/l #	96
40) Benzene	6.044	78	3611280	603.484	ug/l	97
43) Isopropyl Acetate	6.342	43	45286	12.313	ug/l	97
52) Toluene	8.726	92	2496190	639.598	ug/l	94
67) Ethyl Benzene	10.195	91	218276	31.721	ug/l	100
68) m/p-Xylenes	10.305	106	1250114	452.065	ug/l	92
69) o-Xylene	10.646	106	1092900	414.057	ug/l	89
73) Isopropylbenzene	10.963	105	13208	1.785	ug/l	99
75) 1,1,2,2-Tetrachloroethane	11.238	83	42	Below Cal	#	37
78) n-propylbenzene	11.305	91	26028	3.058	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	313255	49.893	ug/l	99
84) 1,2,4-Trimethylbenzene	11.756	105	929993	150.402	ug/l	97
95) Naphthalene	13.786	128	17270752	2431.716	ug/l #	61

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092222\
Data File : VX031600.D
Acq On : 23 Sep 2022 04:07
Operator : JC/MD
Sample : N4731-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SB-1

Quant Time: Sep 23 04:48:18 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091922W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 20 09:45:29 2022
Response via : Initial Calibration

