

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092222\
 Data File : VX031608.D
 Acq On : 23 Sep 2022 07:14
 Operator : JC/MD
 Sample : N4691-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-4

Quant Time: Sep 23 09:20:20 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 20 09:45:29 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	73954	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	121958	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	112284	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	51523	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	74266	54.758	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 109.520%	
35) Dibromofluoromethane	5.385	113	62347	49.452	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 98.900%	
50) Toluene-d8	8.653	98	238284	49.040	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 98.080%	
62) 4-Bromofluorobenzene	11.079	95	72693	44.167	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 88.340%	
Target Compounds						
75) 1,1,2,2-Tetrachloroethane	11.085	83	89	Below Cal	Qvalue #	1

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092222\
Data File : VX031608.D
Acq On : 23 Sep 2022 07:14
Operator : JC/MD
Sample : N4691-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 52 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-4

Quant Time: Sep 23 09:20:20 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091922W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 20 09:45:29 2022
Response via : Initial Calibration

