

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092222\
 Data File : VX031609.D
 Acq On : 23 Sep 2022 07:37
 Operator : JC/MD
 Sample : N4691-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 53 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-SAMPLE-MANAG

Quant Time: Sep 23 09:20:57 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 20 09:45:29 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	72913	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	121117	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	111135	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	50080	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	71302	53.295	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	106.580%
35) Dibromofluoromethane	5.391	113	62039	49.550	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.100%
50) Toluene-d8	8.647	98	235241	48.750	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.500%
62) 4-Bromofluorobenzene	11.079	95	71419	43.694	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.380%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092222\
Data File : VX031609.D
Acq On : 23 Sep 2022 07:37
Operator : JC/MD
Sample : N4691-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 53 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SAMPLE-MANAG

Quant Time: Sep 23 09:20:57 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091922W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 20 09:45:29 2022
Response via : Initial Calibration

